

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

Instrument ID: ECD_S Calibration Date(s): 11/06/2024 11/06/2024
Calibration Times: 09:48 11:24

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028253.D</u>	CF 500 =	<u>PS028254.D</u>		
CF 750 =		<u>PS028255.D</u>	CF 1000 =	<u>PS028256.D</u>	CF 1500 =	<u>PS028257.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	18381400000	17393400000	17071400000	16437600000	15844400000	17025600000	6
2,4-D	3499520000	3198390000	3124710000	3013570000	2944350000	3156110000	7
2,4-DCAA	2911600000	2530960000	2546610000	2432090000	2385800000	2561410000	8

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ENTA05
Lab Code: CHEM **Case No.:** P4660 **SAS No.:** P4660 **SDG NO.:** P4660
Instrument ID: ECD_S **Calibration Date(s):** 11/06/2024 11/06/2024
Calibration Times: 09:48 11:24

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028253.D</u>	CF 500 =	<u>PS028254.D</u>		
CF 750 =		<u>PS028255.D</u>	CF 1000 =	<u>PS028256.D</u>	CF 1500 =	<u>PS028257.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	12950900000	12612000000	12578800000	12192000000	11879700000	12442700000	3
2,4-D	2348170000	2233230000	2226180000	2175820000	2167290000	2230140000	3
2,4-DCAA	1855960000	1741220000	1731630000	1690270000	1689820000	1741780000	4