

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

Continuing Calib Date: 11/06/2024 Initial Calibration Date(s): 11/06/2024 11/06/2024

Continuing Calib Time: 23:03 Initial Calibration Time(s): 09:48 11:24

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.26	7.26	7.16	7.36	0.00
2,4-D	8.39	8.39	8.29	8.49	0.00
2,4,5-TP(Silvex)	9.27	9.28	9.18	9.38	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM **Case No.:** P4660 **SAS No.:** P4660 **SDG NO.:** P4660

Continuing Calib Date: 11/06/2024 **Initial Calibration Date(s):** 11/06/2024 11/06/2024

Continuing Calib Time: 23:03 **Initial Calibration Time(s):** 09:48 11:24

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.76	7.77	7.67	7.87	0.01
2,4-D	9.01	9.02	8.92	9.12	0.01
2,4,5-TP(Silvex)	9.92	9.92	9.82	10.02	0.00