

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL01 Date Analyzed: 11/07/2024

Lab Sample No.: HSTDCCC750 Data File : PS028309.D Time Analyzed: 09:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.269	9.177	9.377	775.150	712.500	8.8
2,4-D	8.387	8.294	8.494	754.140	705.000	7.0
2,4-DCAA	7.256	7.162	7.362	795.060	750.000	6.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL01 Date Analyzed: 11/07/2024

Lab Sample No.: HSTDCCC750 Data File : PS028309.D Time Analyzed: 09:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.917	9.823	10.023	695.710	712.500	-2.4
2,4-D	9.010	8.915	9.115	678.350	705.000	-3.8
2,4-DCAA	7.761	7.665	7.865	740.370	750.000	-1.3