

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL02 Date Analyzed: 11/07/2024

Lab Sample No.: HSTDCCC750 Data File : PS028320.D Time Analyzed: 14:16

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.268	9.177	9.377	786.780	712.500	10.4
2,4-D	8.386	8.294	8.494	760.070	705.000	7.8
2,4-DCAA	7.256	7.162	7.362	800.400	750.000	6.7

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL02 Date Analyzed: 11/07/2024

Lab Sample No.: HSTDCCC750 Data File : PS028320.D Time Analyzed: 14:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.917	9.823	10.023	687.890	712.500	-3.5
2,4-D	9.010	8.915	9.115	670.180	705.000	-4.9
2,4-DCAA	7.761	7.665	7.865	730.880	750.000	-2.5