

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110624

Review By	Abdul	Review On	11/8/2024 11:03:42 AM
Supervise By	Ankita	Supervise On	11/8/2024 11:06:06 AM
SubDirectory	PS110624	HP Acquire Method	HP Processing Method ps110624 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC	PP23462		
Internal Standard/PEM			
ICV/I.BLK	PP23469		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028251.D	06 Nov 2024 09:00		AR\AJ	Ok
2	I.BLK	I.BLK	PS028252.D	06 Nov 2024 09:24		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS028253.D	06 Nov 2024 09:48		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS028254.D	06 Nov 2024 10:12		AR\AJ	Ok,M
5	HSTDICC750	HSTDICC750	PS028255.D	06 Nov 2024 10:36		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS028256.D	06 Nov 2024 11:00		AR\AJ	Ok,M
7	HSTDICC1500	HSTDICC1500	PS028257.D	06 Nov 2024 11:24		AR\AJ	Ok
8	HSTDICV750	ICVPS110624	PS028258.D	06 Nov 2024 11:48		AR\AJ	Ok,M
9	I.BLK	I.BLK	PS028259.D	06 Nov 2024 12:14		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS028260.D	06 Nov 2024 12:38		AR\AJ	Ok
11	P4643-09RE	TP-9RE	PS028261.D	06 Nov 2024 13:02	Surrogate fail	AR\AJ	Confirms
12	PB164519BL	PB164519BL	PS028262.D	06 Nov 2024 13:26		AR\AJ	Ok,M
13	PB164519BS	PB164519BS	PS028263.D	06 Nov 2024 13:50		AR\AJ	Ok
14	PB164393TB	PB164393TB	PS028264.D	06 Nov 2024 14:14	Surrogate Fail	AR\AJ	Not Ok
15	P4578-06	WB-305-BOT	PS028265.D	06 Nov 2024 15:23		AR\AJ	Ok,M
16	P4682-01	BELL-SHOP-RAGS	PS028266.D	06 Nov 2024 15:50		AR\AJ	Ok,M
17	P4667-01	BP-F-6	PS028267.D	06 Nov 2024 16:14		AR\AJ	Ok,M
18	P4667-05	BP-F-5	PS028268.D	06 Nov 2024 16:39		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110624

Review By	Abdul	Review On	11/8/2024 11:03:42 AM
Supervise By	Ankita	Supervise On	11/8/2024 11:06:06 AM
SubDirectory	PS110624	HP Acquire Method	HP Processing Method ps110624 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC	PP23462		
Internal Standard/PEM			
ICV/I.BLK	PP23469		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P4667-09	BP-F-10	PS028269.D	06 Nov 2024 17:03		AR\AJ	Ok,M
20	P4667-13	BP-F-7	PS028270.D	06 Nov 2024 17:27		AR\AJ	Ok,M
21	I.BLK	I.BLK	PS028271.D	06 Nov 2024 17:51		AR\AJ	Ok,M
22	HSTDCCC750	HSTDCCC750	PS028272.D	06 Nov 2024 18:15		AR\AJ	Ok,M
23	P4679-01	MH-1	PS028273.D	06 Nov 2024 18:39		AR\AJ	Ok,M
24	P4680-01	BP-F26	PS028274.D	06 Nov 2024 19:03		AR\AJ	Ok,M
25	P4680-05	BP-F25	PS028275.D	06 Nov 2024 19:27		AR\AJ	Ok,M
26	P4695-01	Z-01	PS028276.D	06 Nov 2024 19:51		AR\AJ	Ok,M
27	P4693-01	BP-G5-WC	PS028277.D	06 Nov 2024 20:15		AR\AJ	Ok,M
28	P4693-05	BP-G4-WC	PS028278.D	06 Nov 2024 20:39		AR\AJ	Ok,M
29	P4711-01	CF-613-COMP-16	PS028279.D	06 Nov 2024 21:03		AR\AJ	Ok,M
30	P4711-06	CF-613-COMP-17	PS028280.D	06 Nov 2024 21:27		AR\AJ	Ok,M
31	P4719-01	BAYAVE-STOCKPILE	PS028281.D	06 Nov 2024 21:51		AR\AJ	Ok,M
32	PB164703BL	PB164703BL	PS028282.D	06 Nov 2024 22:15		AR\AJ	Ok,M
33	I.BLK	I.BLK	PS028283.D	06 Nov 2024 22:39		AR\AJ	Ok,M
34	HSTDCCC750	HSTDCCC750	PS028284.D	06 Nov 2024 23:03		AR\AJ	Ok,M
35	PB164703BS	PB164703BS	PS028285.D	06 Nov 2024 23:51		AR\AJ	Ok
36	PB164494BS	PB164494BS	PS028286.D	07 Nov 2024 00:15		AR\AJ	Ok,M
37	PB164494BSD	PB164494BSD	PS028287.D	07 Nov 2024 00:39		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110624

Review By	Abdul	Review On	11/8/2024 11:03:42 AM
Supervise By	Ankita	Supervise On	11/8/2024 11:06:06 AM
SubDirectory	PS110624	HP Acquire Method	HP Processing Method ps110624 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC	PP23462		
Internal Standard/PEM	PP23469		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	PB164578BS	PB164578BS	PS028288.D	07 Nov 2024 01:04		AR\AJ	Ok,M
39	P4694-01	Z-03A	PS028289.D	07 Nov 2024 01:28		AR\AJ	Ok,M
40	P4694-05	Z-04	PS028290.D	07 Nov 2024 01:52		AR\AJ	Ok,M
41	P4578-06MS	WB-305-BOTMS	PS028291.D	07 Nov 2024 02:16	Some compound recovery fail	AR\AJ	Ok,M
42	P4578-06MSD	WB-305-BOTMSD	PS028292.D	07 Nov 2024 02:40	Some compound recovery fail	AR\AJ	Ok,M
43	P4578-08	WB-305-BOT-1	PS028293.D	07 Nov 2024 03:04		AR\AJ	Ok,M
44	PB164559BL	PB164559BL	PS028294.D	07 Nov 2024 03:28		AR\AJ	Ok,M
45	PB164559BS	PB164559BS	PS028295.D	07 Nov 2024 03:52		AR\AJ	Ok,M
46	I.BLK	I.BLK	PS028296.D	07 Nov 2024 04:16		AR\AJ	Ok
47	HSTDCCC750	HSTDCCC750	PS028297.D	07 Nov 2024 04:40		AR\AJ	Ok,M
48	P4601-19	C0PI0	PS028298.D	07 Nov 2024 05:28		AR\AJ	Ok,M
49	P4601-20	C0PI2	PS028299.D	07 Nov 2024 05:52		AR\AJ	Ok,M
50	P4601-21	C0PI6	PS028300.D	07 Nov 2024 06:16		AR\AJ	Ok,M
51	P4601-22	C0PI8	PS028301.D	07 Nov 2024 06:40		AR\AJ	Ok,M
52	P4601-23	C0PI9	PS028302.D	07 Nov 2024 07:04		AR\AJ	Ok,M
53	P4601-24	CC0P1	PS028303.D	07 Nov 2024 07:28		AR\AJ	Ok,M
54	P4601-25	CC0P3	PS028304.D	07 Nov 2024 07:52		AR\AJ	Ok,M
55	P4601-26	CC0P5	PS028305.D	07 Nov 2024 08:16		AR\AJ	Ok,M
56	P4601-27	CC0P7	PS028306.D	07 Nov 2024 08:40		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110624

Review By	Abdul	Review On	11/8/2024 11:03:42 AM
Supervise By	Ankita	Supervise On	11/8/2024 11:06:06 AM
SubDirectory	PS110624	HP Acquire Method	HP Processing Method ps110624 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC	PP23462		
Internal Standard/PEM			
ICV/I.BLK	PP23469		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

57	P4601-28	CC0P9	PS028307.D	07 Nov 2024 09:04		AR\AJ	Ok,M
58	I.BLK	I.BLK	PS028308.D	07 Nov 2024 09:28		AR\AJ	Ok
59	HSTDCCC750	HSTDCCC750	PS028309.D	07 Nov 2024 09:52		AR\AJ	Ok,M
60	P4660-03	WC-TA2-01-C	PS028310.D	07 Nov 2024 10:16		AR\AJ	Ok
61	P4660-07	WC-WOOD-01-C	PS028311.D	07 Nov 2024 10:40		AR\AJ	Ok,M
62	P4660-11	WC-CONCRETE-01-C	PS028312.D	07 Nov 2024 11:04		AR\AJ	Ok,M
63	P4660-03MS	WC-TA2-01-CMS	PS028313.D	07 Nov 2024 11:28		AR\AJ	Ok,M
64	P4660-03MSD	WC-TA2-01-CMSD	PS028314.D	07 Nov 2024 11:52		AR\AJ	Ok,M
65	PB164752BL	PB164752BL	PS028315.D	07 Nov 2024 12:16		AR\AJ	Ok,M
66	PB164752BS	PB164752BS	PS028316.D	07 Nov 2024 12:40		AR\AJ	Ok
67	PB164560TB	PB164560TB	PS028317.D	07 Nov 2024 13:04		AR\AJ	Ok,M
68	P4368-07	LOD-MDL-WATER-01-C	PS028318.D	07 Nov 2024 13:28		AR\AJ	Ok,M
69	I.BLK	I.BLK	PS028319.D	07 Nov 2024 13:52		AR\AJ	Ok
70	HSTDCCC750	HSTDCCC750	PS028320.D	07 Nov 2024 14:16		AR\AJ	Ok,M
71	P4601-29	CC0Q1	PS028321.D	07 Nov 2024 14:40	Closing CCC Fail	AR\AJ	Not Ok
72	P4601-30	CC0Q6	PS028322.D	07 Nov 2024 15:05	Closing CCC Fail	AR\AJ	Not Ok
73	P4601-31	CC0Q8	PS028323.D	07 Nov 2024 15:29	Closing CCC Fail	AR\AJ	Not Ok
74	P4601-32	CC0R3	PS028324.D	07 Nov 2024 15:53	Closing CCC Fail	AR\AJ	Not Ok
75	P4601-33	CC0R4	PS028325.D	07 Nov 2024 16:17	Closing CCC Fail	AR\AJ	Not Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110624

Review By	Abdul	Review On	11/8/2024 11:03:42 AM
Supervise By	Ankita	Supervise On	11/8/2024 11:06:06 AM
SubDirectory	PS110624	HP Acquire Method	HP Processing Method ps110624 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC Internal Standard/PEM	PP23462		
ICV/I.BLK Surrogate Standard	PP23469		
MS/MSD Standard			
LCS Standard			

76	P4601-34	CC0R5	PS028326.D	07 Nov 2024 16:41	Closing CCC Fail	AR\AJ	Not Ok
77	P4601-35	CC0R6	PS028327.D	07 Nov 2024 17:05	Closing CCC Fail	AR\AJ	Not Ok
78	P4601-36	CC0R7	PS028328.D	07 Nov 2024 17:29	Closing CCC Fail	AR\AJ	Not Ok
79	P4601-37MS	CC0R7MS	PS028329.D	07 Nov 2024 17:53	Closing CCC Fail	AR\AJ	Not Ok
80	P4601-37MSD	CC0R7MSD	PS028330.D	07 Nov 2024 18:17	Closing CCC Fail	AR\AJ	Not Ok
81	I.BLK	I.BLK	PS028331.D	07 Nov 2024 18:41	Closing CCC Fail	AR\AJ	Not Ok
82	HSTDCCC750	HSTDCCC750	PS028332.D	07 Nov 2024 19:05	CCC Fail	AR\AJ	Not Ok

M : Manual Integration