

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

Continuing Calib Date: 11/08/2024 Initial Calibration Date(s): 10/28/2024 10/28/2024

Continuing Calib Time: 12:14 Initial Calibration Time(s): 14:43 15:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.07	9.05	8.95	9.15	-0.02
Tetrachloro-m-xylene	3.55	3.54	3.44	3.64	-0.01
gamma-BHC (Lindane)	4.34	4.33	4.23	4.43	-0.01
Heptachlor	4.92	4.92	4.82	5.02	0.00
Heptachlor epoxide	5.69	5.68	5.58	5.78	-0.01
Endrin	6.58	6.57	6.47	6.67	-0.01
Methoxychlor	7.51	7.50	7.40	7.60	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

Continuing Calib Date: 11/08/2024 Initial Calibration Date(s): 10/28/2024 10/28/2024

Continuing Calib Time: 12:14 Initial Calibration Time(s): 14:43 15:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.92	7.92	7.82	8.02	0.00
Tetrachloro-m-xylene	2.78	2.78	2.68	2.88	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.95	3.95	3.85	4.05	0.00
Heptachlor epoxide	4.74	4.73	4.63	4.83	-0.01
Endrin	5.65	5.64	5.54	5.74	-0.01
Methoxychlor	6.62	6.62	6.52	6.72	0.00