

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/07/2024

Lab Sample No.: PSTDCCC050 Data File : PL092888.D Time Analyzed: 11:21

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Decachlorobiphenyl   | 9.059 | 8.954     | 9.154 | 46.020             | 50.000            | -8.0 |
| Endrin               | 6.577 | 6.474     | 6.674 | 45.550             | 50.000            | -8.9 |
| gamma-BHC (Lindane)  | 4.330 | 4.228     | 4.428 | 48.790             | 50.000            | -2.4 |
| Heptachlor           | 4.919 | 4.817     | 5.017 | 47.760             | 50.000            | -4.5 |
| Heptachlor epoxide   | 5.686 | 5.584     | 5.784 | 47.690             | 50.000            | -4.6 |
| Methoxychlor         | 7.502 | 7.399     | 7.599 | 47.810             | 50.000            | -4.4 |
| Tetrachloro-m-xylene | 3.542 | 3.440     | 3.640 | 48.470             | 50.000            | -3.1 |

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/07/2024

Lab Sample No.: PSTDCCC050 Data File : PL092888.D Time Analyzed: 11:21

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Decachlorobiphenyl   | 7.917 | 7.816     | 8.016 | 46.840             | 50.000            | -6.3 |
| Endrin               | 5.642 | 5.541     | 5.741 | 51.530             | 50.000            | 3.1  |
| gamma-BHC (Lindane)  | 3.611 | 3.511     | 3.711 | 50.380             | 50.000            | 0.8  |
| Heptachlor           | 3.950 | 3.849     | 4.049 | 50.240             | 50.000            | 0.5  |
| Heptachlor epoxide   | 4.732 | 4.632     | 4.832 | 50.240             | 50.000            | 0.5  |
| Methoxychlor         | 6.615 | 6.515     | 6.715 | 49.970             | 50.000            | -0.1 |
| Tetrachloro-m-xylene | 2.778 | 2.678     | 2.878 | 49.610             | 50.000            | -0.8 |