

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/07/2024

Lab Sample No.: PSTDCCC050 Data File : PL092904.D Time Analyzed: 15:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.059	8.954	9.154	45.750	50.000	-8.5
Endrin	6.577	6.474	6.674	42.920	50.000	-14.2
gamma-BHC (Lindane)	4.330	4.228	4.428	48.200	50.000	-3.6
Heptachlor	4.919	4.817	5.017	46.410	50.000	-7.2
Heptachlor epoxide	5.687	5.584	5.784	45.780	50.000	-8.4
Methoxychlor	7.503	7.399	7.599	43.270	50.000	-13.5
Tetrachloro-m-xylene	3.542	3.440	3.640	47.550	50.000	-4.9

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/07/2024

Lab Sample No.: PSTDCCC050 Data File : PL092904.D Time Analyzed: 15:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.917	7.816	8.016	40.370	50.000	-19.3
Endrin	5.642	5.541	5.741	46.330	50.000	-7.3
gamma-BHC (Lindane)	3.611	3.511	3.711	49.880	50.000	-0.2
Heptachlor	3.950	3.849	4.049	49.860	50.000	-0.3
Heptachlor epoxide	4.732	4.632	4.832	47.410	50.000	-5.2
Methoxychlor	6.616	6.515	6.715	44.210	50.000	-11.6
Tetrachloro-m-xylene	2.778	2.678	2.878	50.400	50.000	0.8