

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL03 Date Analyzed: 11/08/2024

Lab Sample No.: PSTDCCC050 Data File : PL092920.D Time Analyzed: 12:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.068	8.954	9.154	47.820	50.000	-4.4
Endrin	6.583	6.474	6.674	46.140	50.000	-7.7
gamma-BHC (Lindane)	4.337	4.228	4.428	50.230	50.000	0.5
Heptachlor	4.924	4.817	5.017	49.050	50.000	-1.9
Heptachlor epoxide	5.694	5.584	5.784	49.150	50.000	-1.7
Methoxychlor	7.510	7.399	7.599	46.690	50.000	-6.6
Tetrachloro-m-xylene	3.548	3.440	3.640	50.310	50.000	0.6

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL03 Date Analyzed: 11/08/2024

Lab Sample No.: PSTDCCC050 Data File : PL092920.D Time Analyzed: 12:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.921	7.816	8.016	49.630	50.000	-0.7
Endrin	5.646	5.541	5.741	53.050	50.000	6.1
gamma-BHC (Lindane)	3.613	3.511	3.711	53.370	50.000	6.7
Heptachlor	3.952	3.849	4.049	52.180	50.000	4.4
Heptachlor epoxide	4.735	4.632	4.832	53.410	50.000	6.8
Methoxychlor	6.619	6.515	6.715	50.650	50.000	1.3
Tetrachloro-m-xylene	2.780	2.678	2.878	52.400	50.000	4.8