

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL04 Date Analyzed: 11/08/2024

Lab Sample No.: PSTDCCC050 Data File : PL092939.D Time Analyzed: 17:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.060	8.954	9.154	42.510	50.000	-15.0
Endrin	6.578	6.474	6.674	41.150	50.000	-17.7
gamma-BHC (Lindane)	4.330	4.228	4.428	49.260	50.000	-1.5
Heptachlor	4.919	4.817	5.017	45.670	50.000	-8.7
Heptachlor epoxide	5.687	5.584	5.784	45.730	50.000	-8.5
Methoxychlor	7.503	7.399	7.599	40.630	50.000	-18.7
Tetrachloro-m-xylene	3.542	3.440	3.640	49.900	50.000	-0.2

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL04 Date Analyzed: 11/08/2024

Lab Sample No.: PSTDCCC050 Data File : PL092939.D Time Analyzed: 17:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.919	7.816	8.016	43.240	50.000	-13.5
Endrin	5.643	5.541	5.741	48.460	50.000	-3.1
gamma-BHC (Lindane)	3.612	3.511	3.711	53.000	50.000	6.0
Heptachlor	3.951	3.849	4.049	50.120	50.000	0.2
Heptachlor epoxide	4.733	4.632	4.832	52.110	50.000	4.2
Methoxychlor	6.617	6.515	6.715	43.800	50.000	-12.4
Tetrachloro-m-xylene	2.779	2.678	2.878	52.590	50.000	5.2