

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL05 Date Analyzed: 11/11/2024

Lab Sample No.: PSTDCCC050 Data File : PL092943.D Time Analyzed: 11:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.061	8.954	9.154	46.430	50.000	-7.1
Endrin	6.579	6.474	6.674	43.850	50.000	-12.3
gamma-BHC (Lindane)	4.331	4.228	4.428	49.900	50.000	-0.2
Heptachlor	4.920	4.817	5.017	47.730	50.000	-4.5
Heptachlor epoxide	5.688	5.584	5.784	48.400	50.000	-3.2
Methoxychlor	7.504	7.399	7.599	44.210	50.000	-11.6
Tetrachloro-m-xylene	3.542	3.440	3.640	49.960	50.000	-0.1

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL05 Date Analyzed: 11/11/2024

Lab Sample No.: PSTDCCC050 Data File : PL092943.D Time Analyzed: 11:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.919	7.816	8.016	46.770	50.000	-6.5
Endrin	5.643	5.541	5.741	50.950	50.000	1.9
gamma-BHC (Lindane)	3.611	3.511	3.711	52.560	50.000	5.1
Heptachlor	3.951	3.849	4.049	50.950	50.000	1.9
Heptachlor epoxide	4.733	4.632	4.832	52.470	50.000	4.9
Methoxychlor	6.617	6.515	6.715	46.930	50.000	-6.1
Tetrachloro-m-xylene	2.778	2.678	2.878	51.750	50.000	3.5