

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE102824

Review By	yogesh	Review On	10/29/2024 4:57:44 AM
Supervise By	mohammad	Supervise On	10/30/2024 2:48:38 AM
SubDirectory	BE102824	HP Acquire Method	BNA_E
		HP Processing Method	be102824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12323,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BE101388.D	28 Oct 2024 10:51		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BE101389.D	28 Oct 2024 11:44		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BE101390.D	28 Oct 2024 12:23	Compound #32,54,56,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BE101391.D	28 Oct 2024 12:59		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BE101392.D	28 Oct 2024 13:35	Compound #32,54 Kept on LR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BE101393.D	28 Oct 2024 14:11	The Calibration is Good For 8270 DOD Except com#77 and Not good for 625.1 Method	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BE101394.D	28 Oct 2024 14:49	Com#77(Benzidine) Failed in the calibration for both DOD and NON-DOD	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BE101395.D	28 Oct 2024 15:25	Compound #69 removed from 60ppm	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BE101396.D	28 Oct 2024 16:01	Compound #9,69,79 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBE102824	BE101397.D	28 Oct 2024 16:37	ICV Fail for Com#77 for Both DOD and NON-DOD	RC/JU	Ok,M
11	PB164444BL	PB164444BL	BE101398.D	28 Oct 2024 17:16		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110524

Review By	yogesh	Review On	11/6/2024 3:08:53 AM
Supervise By	mohammad	Supervise On	11/6/2024 3:16:26 AM
SubDirectory	BE110524	HP Acquire Method	BNA_E
		HP Processing Method	be102824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12324,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BE101471.D	5 Nov 2024 10:12		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BE101472.D	5 Nov 2024 10:48		RC/JU	Ok,M
3	PB164639BL	PB164639BL	BE101473.D	5 Nov 2024 11:33		RC/JU	Ok
4	PB164639BS	PB164639BS	BE101474.D	5 Nov 2024 12:09		RC/JU	Ok,M
5	P4680-01	BP-F26	BE101475.D	5 Nov 2024 12:44		RC/JU	Ok
6	P4675-02	COMP-2	BE101476.D	5 Nov 2024 13:20		RC/JU	Ok,M
7	P4675-05	COMP-5	BE101477.D	5 Nov 2024 13:56		RC/JU	Ok
8	P4675-06	COMP-6	BE101478.D	5 Nov 2024 14:32		RC/JU	Ok
9	P4667-05	BP-F-5	BE101479.D	5 Nov 2024 15:08		RC/JU	Ok
10	P4640-04	MH-3	BE101480.D	5 Nov 2024 15:43		RC/JU	Ok
11	P4640-04MS	MH-3MS	BE101481.D	5 Nov 2024 16:16		RC/JU	Ok,M
12	P4640-04MSD	MH-3MSD	BE101482.D	5 Nov 2024 16:52		RC/JU	Ok,M
13	P4643-04	BP-F9-ADDITIONAL	BE101483.D	5 Nov 2024 17:28		RC/JU	Ok
14	P4643-08	BP-F8	BE101484.D	5 Nov 2024 18:04		RC/JU	Ok
15	P4643-12	TP-9	BE101485.D	5 Nov 2024 18:39		RC/JU	Ok
16	P4682-02	BELL-SHOP-RAGS	BE101486.D	5 Nov 2024 19:15		RC/JU	Ok
17	P4645-04	Z-02-WC	BE101487.D	5 Nov 2024 19:51		RC/JU	Ok
18	P4660-03	WC-TA2-01-C	BE101488.D	5 Nov 2024 20:27		RC/JU	Ok

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110524

Review By	yogesh	Review On	11/6/2024 3:08:53 AM		
Supervise By	mohammad	Supervise On	11/6/2024 3:16:26 AM		
SubDirectory	BE110524	HP Acquire Method	BNA_E	HP Processing Method	be102824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12324,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	P4682-01	BELL-SHOP-RAGS	BE101489.D	5 Nov 2024 21:03	Internal Standrad Fail	RC/JU	Ok,M
20	P4685-01	OK-01-11012024	BE101490.D	5 Nov 2024 21:38		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110624

Review By	Jagrut	Review On	11/7/2024 11:03:46 AM
Supervise By	mohammad	Supervise On	11/8/2024 3:53:54 AM
SubDirectory	BE110624	HP Acquire Method	BNA_E
		HP Processing Method	be110624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12324,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BE101491.D	6 Nov 2024 11:40		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BE101492.D	6 Nov 2024 13:15	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BE101493.D	6 Nov 2024 13:51		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BE101494.D	6 Nov 2024 14:27	Compound #32,54 removed from 5ppm	RC/JU	Ok,M
5	SSTDICC010	SSTDICC010	BE101495.D	6 Nov 2024 15:03		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BE101496.D	6 Nov 2024 15:39	Compound #32 Kept on LR	RC/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BE101497.D	6 Nov 2024 16:14	The Calibration is Good For 8270 DOD Except com#77 and will NOT be used for 625.1 Method	RC/JU	Ok,M
8	SSTDICC050	SSTDICC050	BE101498.D	6 Nov 2024 16:50	Com#77(Benzidine) Failed in the calibration for both DOD and NON-DOD	RC/JU	Ok,M
9	SSTDICC060	SSTDICC060	BE101499.D	6 Nov 2024 17:26		RC/JU	Ok,M
10	SSTDICC080	SSTDICC080	BE101500.D	6 Nov 2024 18:02		RC/JU	Ok,M
11	SSTDICV040	ICVBE110624	BE101501.D	6 Nov 2024 18:41		RC/JU	Ok,M
12	PB164561BL	PB164561BL	BE101502.D	6 Nov 2024 19:16		RC/JU	Ok
13	DFTPP	DFTPP	BE101503.D	6 Nov 2024 19:52		RC/JU	Ok
14	SSTDCCC040	SSTDCCC040	BE101504.D	6 Nov 2024 20:28		RC/JU	Ok,M
15	PB164561TB	PB164561TB	BE101505.D	6 Nov 2024 21:04		RC/JU	Ok
16	PB164561BS	PB164561BS	BE101506.D	6 Nov 2024 21:40		RC/JU	Ok,M

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110624

Review By	Jagrut	Review On	11/7/2024 11:03:46 AM		
Supervise By	mohammad	Supervise On	11/8/2024 3:53:54 AM		
SubDirectory	BE110624	HP Acquire Method	BNA_E	HP Processing Method	be110624
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12324,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

17	P4660-11	WC-CONCRETE-01-C	BE101507.D	6 Nov 2024 22:15		RC/JU	Ok
18	P4667-04	BP-F-6	BE101508.D	6 Nov 2024 22:51		RC/JU	Ok
19	P4667-08	BP-F-5	BE101509.D	6 Nov 2024 23:27	Internal Standard and Surrogate fail	RC/JU	ReRun
20	P4667-12	TP-10	BE101510.D	7 Nov 2024 00:03	Internal Standard Fail	RC/JU	ReRun
21	P4667-16	BP-F-7	BE101511.D	7 Nov 2024 00:39		RC/JU	Ok
22	P4659-04	MH-2	BE101512.D	7 Nov 2024 1:14		RC/JU	Ok
23	P4679-04	MH-1	BE101513.D	7 Nov 2024 1:50		RC/JU	Ok
24	P4680-04	BP-F26	BE101514.D	7 Nov 2024 2:26	Internal Standrad Fail	RC/JU	ReRun
25	P4680-08	BP-F25	BE101515.D	7 Nov 2024 3:02	Internal Standrad Fail	RC/JU	ReRun
26	P4660-07	WC-WOOD-01-C	BE101516.D	7 Nov 2024 3:38		RC/JU	Ok,M
27	P4711-05	CF-613-COMP-16	BE101517.D	7 Nov 2024 4:13	Internal Standrad Fail	RC/JU	ReRun
28	P4711-10	CF-613-COMP-17	BE101518.D	7 Nov 2024 4:49	Internal Standrad Fail	RC/JU	ReRun
29	P4711-01	CF-613-COMP-16	BE101519.D	7 Nov 2024 5:25		RC/JU	Ok,M
30	P4711-06	CF-613-COMP-17	BE101520.D	7 Nov 2024 6:01	Need further 5X Dilution	RC/JU	Dilution
31	P4708-01	OR-02-110424	BE101521.D	7 Nov 2024 6:37		RC/JU	Ok,M
32	P4706-01	TR-04-110424	BE101522.D	7 Nov 2024 7:13		RC/JU	Ok,M

M : Manual Integration