

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4660

Client: ENTACT

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P4640-04MSD	Client Sample ID:	MH-3MSD					DataFile:	BE101482.D		
Pyridine	500	0	270	ug/L	54		4		20 (10)	160 (109)	20 (20)
1,4-Dichlorobenzene	500	0	260	ug/L	52	*	0		70 (55)	130 (125)	20 (20)
2-Methylphenol	500	0	440	ug/L	88		10		70 (37)	130 (126)	20 (20)
3+4-Methylphenols	500	0	450	ug/L	90		9		20 (31)	160 (127)	20 (20)
Hexachloroethane	500	0	270	ug/L	54		4		20 (49)	160 (110)	20 (20)
Nitrobenzene	500	0	390	ug/L	78		5		70 (62)	130 (112)	20 (20)
Hexachlorobutadiene	500	0	340	ug/L	68	*	3		70 (52)	130 (125)	20 (20)
2,4,6-Trichlorophenol	500	0	530	ug/L	106		6		70 (78)	130 (112)	20 (20)
2,4,5-Trichlorophenol	500	0	510	ug/L	102		6		70 (71)	130 (111)	20 (20)
2,4-Dinitrotoluene	500	0	480	ug/L	96		4		70 (50)	130 (142)	20 (20)
Hexachlorobenzene	500	0	490	ug/L	98		4		70 (72)	130 (115)	20 (20)
Pentachlorophenol	1000	0	1000	ug/L	100		1		20 (25)	160 (139)	20 (20)