

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/05/2024

Lab File ID: BE101472.D Time Analyzed: 10:48

Instrument ID: BNA\_E GC Column: ZB-GR ID: 0.25 (mm)

|                | IS1 (DCB)<br>AREA # | RT # | IS2 (NPT)<br>AREA # | RT #   | IS3 (ANT)<br>AREA # | RT #  |
|----------------|---------------------|------|---------------------|--------|---------------------|-------|
| 12 HOUR STD    | 77150               | 7.57 | 373279              | 10.34  | 260925              | 14.18 |
| UPPER LIMIT    | 154300              | 8.07 | 746558              | 10.838 | 521850              | 14.68 |
| LOWER LIMIT    | 38575               | 7.07 | 186640              | 9.838  | 130463              | 13.68 |
| EPA SAMPLE NO. |                     |      |                     |        |                     |       |
| 01 MH-3MS      | 74749               | 7.57 | 324674              | 10.34  | 214275              | 14.18 |
| 02 MH-3MSD     | 70903               | 7.57 | 316494              | 10.34  | 211098              | 14.18 |
| 03 WC-TA2-01-C | 57343               | 7.57 | 245737              | 10.34  | 174187              | 14.18 |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.