

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

EPA Sample No.: SSTDICCC040 Date Analyzed: 11/06/2024

Lab File ID: BE101497.D Time Analyzed: 16:14

Instrument ID: BNA_E GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	65428	7.563	303330	10.33	215457	14.17
UPPER LIMIT	130856	8.063	606660	10.831	430914	14.673
LOWER LIMIT	32714	7.063	151665	9.831	107729	13.673
EPA SAMPLE NO.						
01 PB164561BL	49754	7.56	204369	10.33	133519	14.17

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.