

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/06/2024

Lab File ID: BE101504.D Time Analyzed: 20:28

Instrument ID: BNA_E GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	70453	7.559	326960	10.33	230956	14.18
UPPER LIMIT	140906	8.059	653920	10.832	461912	14.675
LOWER LIMIT	35226.5	7.059	163480	9.832	115478	13.675
EPA SAMPLE NO.						
01 PB164561BS	49257	7.56	217361	10.33	139774	14.17
02 PB164561TB	53927	7.57	223555	10.33	146954	14.17
03 WC-CONCRETE-01-C	46975	7.56	200241	10.33	139072	14.17
04 WC-WOOD-01-C	43720	7.56	200694	10.33	141864	14.17

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.