

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660

EPA Sample No.: SSTDICCC040 Date Analyzed: 11/06/2024

Lab File ID: BE101497.D Time Analyzed: 16:14

Instrument ID: BNA_E GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	491742	16.911	550275	21.077	716867	23.357
UPPER LIMIT	983484	17.411	1100550	21.577	1433730	23.857
LOWER LIMIT	245871	16.411	275138	20.577	358434	22.857
EPA SAMPLE NO.						
01 PB164561BL	306761	16.91	392502	21.07	573520	23.35

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.