

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: ENTA05

Lab Code: CHEM

SAS No.: P4660 SDG NO.: P4660

Lab File ID: BE101471.D

DFTPP Injection Date: 11/05/2024

Instrument ID: BNA_E

DFTPP Injection Time: 10:12

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	14
68	Less than 2.0% of mass 69	0.2 (1.5) 1
69	Mass 69 relative abundance	15.1
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	22
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	3.4
275	10.0 - 60.0% of mass 198	17.6
365	Greater than 1% of mass 198	2.9
441	Present, but less than mass 443	16.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.3 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BE101472.D	11/05/2024	10:48
MH-3MS	P4640-04MS	BE101481.D	11/05/2024	16:16
MH-3MSD	P4640-04MSD	BE101482.D	11/05/2024	16:52
WC-TA2-01-C	P4660-03	BE101488.D	11/05/2024	20:27