

Report of Analysis

Client:	ENTACT		Date Collected:		
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:		
Client Sample ID:	VN1104WBS01		SDG No.:	P4660	
Lab Sample ID:	VN1104WBS01		Matrix:	TCLP	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	TCLP VOA	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084648.D	1		11/04/24 11:14	VN110424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	18.1		0.34	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.5		0.26	1.00	ug/L
78-93-3	2-Butanone	93.0		1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.0		0.25	1.00	ug/L
67-66-3	Chloroform	18.6		0.26	1.00	ug/L
71-43-2	Benzene	18.3		0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.8		0.24	1.00	ug/L
79-01-6	Trichloroethene	18.2		0.32	1.00	ug/L
127-18-4	Tetrachloroethene	17.9		0.25	1.00	ug/L
108-90-7	Chlorobenzene	17.3		0.13	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	47.2		70 (74) - 130 (125)	94%	SPK: 50
1868-53-7	Dibromofluoromethane	49.6		70 (75) - 130 (124)	99%	SPK: 50
2037-26-5	Toluene-d8	49.7		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.4		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	191000	8.218			
540-36-3	1,4-Difluorobenzene	316000	9.094			
3114-55-4	Chlorobenzene-d5	287000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	156000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products