

Surrogate Summary

SDG No.: P4660

Client: ENTACT

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4660-01	WC-TA2-01-G	1,2-Dichloroethane-d4	50	50.5	101	70 (74)	130 (125)
		Dibromofluoromethane	50	46.1	92	70 (75)	130 (124)
		Toluene-d8	50	46.1	92	70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.3	91	70 (77)	130 (121)
P4660-05	WC-WOOD-01-G	1,2-Dichloroethane-d4	50	49.3	99	70 (74)	130 (125)
		Dibromofluoromethane	50	48.1	96	70 (75)	130 (124)
		Toluene-d8	50	49.0	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.0	100	70 (77)	130 (121)
P4660-09	WC-CONCRETE-01-G	1,2-Dichloroethane-d4	50	46.5	93	70 (74)	130 (125)
		Dibromofluoromethane	50	46.1	92	70 (75)	130 (124)
		Toluene-d8	50	47.6	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.5	103	70 (77)	130 (121)

Surrogate Summary

SDG No.: P4660

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VN1104WBL01	VN1104WBL01	1,2-Dichloroethane-d4	50	49.3	98	70 (74)	130 (125)
		Dibromofluoromethane	50	49.8	100	70 (75)	130 (124)
		Toluene-d8	50	46.9	94	70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.1	90	70 (77)	130 (121)
VN1104WBS01	VN1104WBS01	1,2-Dichloroethane-d4	50	47.2	94	70 (74)	130 (125)
		Dibromofluoromethane	50	49.6	99	70 (75)	130 (124)
		Toluene-d8	50	49.7	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.4	103	70 (77)	130 (121)
VN1104WBSD0	VN1104WBSD01	1,2-Dichloroethane-d4	50	48.1	96	70 (74)	130 (125)
		Dibromofluoromethane	50	48.0	96	70 (75)	130 (124)
		Toluene-d8	50	49.0	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.9	96	70 (77)	130 (121)

