

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4660

Client: ENTACT

Analytical Method: SW8260-Low

Datafile : VN084649.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1104WBSD01	Vinyl chloride	20	18.6	ug/L	93	2		70 (65)	130 (117)	20 (20)
	1,1-Dichloroethene	20	18.0	ug/L	90	2		70 (74)	130 (110)	20 (20)
	2-Butanone	100	99.4	ug/L	99	6		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	19.2	ug/L	96	1		70 (77)	130 (113)	20 (20)
	Chloroform	20	19.8	ug/L	99	6		70 (79)	130 (113)	20 (20)
	Benzene	20	19.1	ug/L	96	4		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	19.7	ug/L	99	5		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	18.7	ug/L	94	3		70 (77)	130 (113)	20 (20)
	Tetrachloroethene	20	19.4	ug/L	97	7		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	19.0	ug/L	95	10		70 (82)	130 (109)	20 (20)