

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660
 Lab File ID: VN084645.D Date Analyzed: 11/04/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:32
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	188408	8.22	303241	9.10	268457	11.86
UPPER LIMIT	376816	8.718	606482	9.6	536914	12.359
LOWER LIMIT	94204	7.718	151621	8.6	134229	11.359
EPA SAMPLE NO.						
WC-TA2-01-G	171677	8.22	307928	9.10	271085	11.87
WC-WOOD-01-G	170730	8.22	303877	9.10	288279	11.87
WC-CONCRETE-01-G	183096	8.22	318203	9.10	292659	11.87
VN1104WBL01	181983	8.22	315820	9.09	273716	11.86
VN1104WBS01	191080	8.22	315901	9.09	287009	11.87
VN1104WBSD01	179675	8.22	301837	9.10	269118	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.