

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P4660 SAS No.: P4660 SDG NO.: P4660
 Lab File ID: VN084645.D Date Analyzed: 11/04/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:32
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	143756	13.788				
UPPER LIMIT	287512	14.288				
LOWER LIMIT	71878	13.288				
EPA SAMPLE NO.						
WC-TA2-01-G	119210	13.79				
WC-WOOD-01-G	139356	13.79				
WC-CONCRETE-01-G	141822	13.79				
VN1104WBL01	126703	13.79				
VN1104WBS01	156261	13.79				
VN1104WBSD01	133565	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.