

CASE NARRATIVE

Union Paving & Construction Company

Project Name: Rt. 10 Whippany

Project # N/A

Chemtech Project # P4663

Test Name: Herbicide

A. Number of Samples and Date of Receipt:

8 Solid samples were received on 10/31/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Cyanide, EPH_NF, Herbicide, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, PCB, Pesticide-TCL, SVOCMS Group1, TPH GC, Trivalent Chromium and VOCMS Group1. This data package contains results for Herbicide.

C. Analytical Techniques:

The analysis was performed on instrument ECD_S. The front column is RTX-CLPesticides which is 30 meters, 0.32 mm ID, 0.5 um df, Catalog # 11139. The rear column is RTX-CLPesticides2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 11324. The analysis of Herbicides was based on method 8151A and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for CC0R8MS [2,4-DCAA(2) - 51%], CC0R8MSD [2,4-DCAA(2) - 51%], TENNANT-PAD-2-3-STOCKPILE [2,4-DCAA(1) - 56%, 2,4-DCAA(2) - 51%], RES-BUILDING-PAD-NO-11 [2,4-DCAA(1) - 54%, 2,4-DCAA(2) - 49%], RES-BUILDING-PAD-NO-3 [2,4-DCAA(1) - 50%, 2,4-DCAA(2) - 39%], RES-BUILDING-PAD-NO-2 [2 and 4-DCAA(2) - 66%] these compounds did not meet the NJDKQP criteria but met the in-house criteria.

The Retention Times were acceptable for all samples.

The MS {P4602-44MS} with File ID: PS028408.D recoveries met the requirements for all compounds except for 2,4,5-T[27%], 2,4,5-TP(Silvex)[16%], 2,4-D[50%], 2,4-DB[29%], DICAMBA[43%], DICHLORPROP[26%] these compounds did not meet the NJDKQP criteria but met the in-house criteria and Dinoseb[0%] this compound did not meet the NJDKQP criteria and in-house criteria due to sample matrix interference.

The MSD {P4602-45MSD} with File ID: PS028409.D recoveries met the acceptable requirements except for 2,4,5-T[28%], 2,4,5-TP(Silvex)[16%], 2,4-D[50%], 2,4-



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

DB[29%], DICAMBA[43%], DICHLORPROP[26%]these compounds did not meet the NJDKQP criteria but met the in-house criteria and Dinoseb[0%]this compound did not meet the NJDKQP criteria and in-house criteria due to sample matrix interference.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration File ID PS028237.D met the requirements except for 2,4,5-T,2,4,5-TP (Silvex),2,4-D,2,4-DB,2,4-DCAA,DICAMBA,DICHLORPROP,DINOSEB is failing in 1st column but it is passing in 2nd column also associated samples not having any positive hit therefore no corrective action taken.

E. Additional Comments:

Fax and Hardcopy data will not match for all samples as Fax samples analyzed in sequence PS110424 where CCAL having F flag as a corrective action sample reanalyzed and reported in hardcopy and fax data From sequence PS110424 provided in Miscellaneous Section.

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____