

CASE NARRATIVE

Union Paving & Construction Company

Project Name: Rt. 10 Whippany

Project # N/A

Chemtech Project # P4663

Test Name: SVOCMS Group1

A. Number of Samples and Date of Receipt:

8 Solid samples were received on 10/31/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Cyanide, EPH_NF, Herbicide, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, PCB, Pesticide-TCL, SVOCMS Group1, TPH GC, Trivalent Chromium and VOCMS Group1. This data package contains results for SVOCMS Group1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_E using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um dfThe analysis of SVOCMS Group1 was based on method 8270E and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {P4663-08MS} with File ID: BF140209.D recoveries met the requirements for all compounds except for 4,6-Dinitro-2-methylphenol[139%], 4-Chloroaniline[43%], bis(2-Ethylhexyl)phthalate[167%], Di-n-octyl phthalate[133%], Hexachlorocyclopentadiene[164%] these compounds did not meet the NJDKQP criteria but met the in-house criteria and Butylbenzylphthalate[150%] this compound did not meet the NJDKQP criteria and in-house criteria due to matrix interference.

The MSD {P4663-08MSD} with File ID: BF140210.D recoveries met the acceptable requirements except for 4,6-Dinitro-2-methylphenol[150%], 4-Chloroaniline[46%], Di-n-octyl phthalate[144%], Hexachlorocyclopentadiene[169%] these compounds did not meet the NJDKQP criteria but met the in-house criteria and Butylbenzylphthalate[167%], bis(2-Ethylhexyl)phthalate[183%] these compounds did not meet the NJDKQP criteria and in-house criteria due to matrix interference.

The RPD met criteria .

The Blank Spike for {PB164593BS} with File ID: BE101455.D met requirements for all samples except for 3-Nitroaniline[65%] this compound did not meet the NJDKQP criteria but met the in-house criteria and Hexachlorocyclopentadiene[215%] this compound did not meet the NJDKQP criteria and in-house criteria but no positive hits in associated samples therefore no corrective action taken.

The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 20% in the Initial Calibration (Method 8270-BE102824.M) for 2,4-Dinitrophenol, this compound is passing on Linear Regression.

The % RSD is greater than 20% in the Initial Calibration (Method 8270-BF101824.M) for 2,4-Dinitrophenol, this compound is passing on Linear Regression.

The Continuous Calibration File ID BF140200.D met the requirements except for 2,4-Dinitrophenol, 2,4-Dinitrotoluene, 2-Nitrophenol, 4,6-Dinitro-2-methylphenol, 4-Nitroaniline, 4-Nitrophenol, Bis(2-ethylhexyl)phthalate, Butylbenzylphthalate and Di-n-octyl phthalate but no positive hits in associated samples therefore no corrective action taken.

The Tuning criteria met requirements.

E. Additional Comments:

For sample # TENNANT-PAD-2-3-STOCKPILE some compounds below Method detection limits, therefore it is not reported as Hit in Form-1.

The Form 6 is not included in the data package because the Initial Calibration was performed using 8 points.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____