

Initial and Continuing Calibration Blank Summary

Client: Union Paving & Construction Company

SDG No.: P4663

Project: Rt. 10 Whippany

RunNo.: LB133276

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	11/04/2024
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	11/04/2024
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	11/04/2024
Sample ID: CCB3 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	11/04/2024
Sample ID: CCB4 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	11/04/2024

Initial and Continuing Calibration Blank Summary

Client: Union Paving & Construction Company

SDG No.: P4663

Project: Rt. 10 Whippany

RunNo.: LB133277

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	11/04/2024
Sample ID: CCB1 Cyanide	mg/L	0.0017	0.0025	J	0.00099	0.005	11/04/2024
Sample ID: CCB2 Cyanide	mg/L	0.0011	0.0025	J	0.00099	0.005	11/04/2024
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	11/04/2024