

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133275

Review By	mohan	Review On	11/6/2024 1:07:46 AM
Supervise By	jaswal	Supervise On	11/6/2024 1:08:59 AM

STD. NAME	STD REF.#
ICAL Standard	MP83017,MP83018,MP83019,MP83020,MP83021,MP83022
ICV Standard	MP83023
CCV Standard	MP83025
ICSA Standard	
CRI Standard	MP83027
LCS Standard	
Chk Standard	MP83024,MP83026,,,MP83028,MP83030

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/04/24 13:47		Mohan	OK
2	S0.2	S0.2	CAL2	11/04/24 13:49		Mohan	OK
3	S2.5	S2.5	CAL3	11/04/24 13:51		Mohan	OK
4	S5	S5	CAL4	11/04/24 13:53		Mohan	OK
5	S7.5	S7.5	CAL5	11/04/24 13:56		Mohan	OK
6	S10	S10	CAL6	11/04/24 14:08		Mohan	OK
7	ICV68	ICV68	ICV	11/04/24 14:11		Mohan	OK
8	ICB68	ICB68	ICB	11/04/24 14:16		Mohan	OK
9	CCV22	CCV22	CCV	11/04/24 14:18		Mohan	OK
10	CCB22	CCB22	CCB	11/04/24 14:23		Mohan	OK
11	CRA	CRA	CRDL	11/04/24 14:25		Mohan	OK
12	HighStd	HighStd	HIGH STD	11/04/24 14:27		Mohan	OK
13	ChkStd	ChkStd	SAM	11/04/24 14:30		Mohan	OK
14	PB164668BL	PB164668BL	MB	11/04/24 14:32		Mohan	OK
15	PB164668BS	PB164668BS	LCS	11/04/24 14:34		Mohan	OK
16	P4663-02MS	TENNANT-PAD-2-3-S	MS	11/04/24 14:41		Mohan	OK
17	P4663-02MSD	TENNANT-PAD-2-3-S	MSD	11/04/24 14:43		Mohan	OK
18	P4663-04	RES-BUILDING-PAD-	SAM	11/04/24 14:46		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133275

Review By	mohan	Review On	11/6/2024 1:07:46 AM
Supervise By	jaswal	Supervise On	11/6/2024 1:08:59 AM

STD. NAME	STD REF.#
ICAL Standard	MP83017,MP83018,MP83019,MP83020,MP83021,MP83022
ICV Standard	MP83023
CCV Standard	MP83025
ICSA Standard	
CRI Standard	MP83027
LCS Standard	
Chk Standard	MP83024,MP83026,,,MP83028,MP83030

19	CCV23	CCV23	CCV	11/04/24 14:48		Mohan	OK
20	CCB23	CCB23	CCB	11/04/24 14:53		Mohan	OK
21	P4663-06	RES-BUILDING-PAD-	SAM	11/04/24 14:55		Mohan	OK
22	P4663-08	RES-BUILDING-PAD-	SAM	11/04/24 14:57		Mohan	OK
23	P4667-01	BP-F-6	SAM	11/04/24 15:00		Mohan	OK
24	P4667-05	BP-F-5	SAM	11/04/24 15:02		Mohan	OK
25	P4667-09	TP-10	SAM	11/04/24 15:04		Mohan	OK
26	P4667-13	BP-F-7	SAM	11/04/24 15:06		Mohan	OK
27	P4675-01	COMP-1	SAM	11/04/24 15:09		Mohan	OK
28	P4675-02	COMP-2	SAM	11/04/24 15:11		Mohan	OK
29	P4675-03	COMP-3	SAM	11/04/24 15:13		Mohan	OK
30	P4675-04	COMP-4	SAM	11/04/24 15:16		Mohan	OK
31	CCV24	CCV24	CCV	11/04/24 15:18		Mohan	OK
32	CCB24	CCB24	CCB	11/04/24 15:20		Mohan	OK
33	P4675-05	COMP-5	SAM	11/04/24 15:22		Mohan	OK
34	P4675-06	COMP-6	SAM	11/04/24 15:25		Mohan	OK
35	P4679-01	MH-1	SAM	11/04/24 15:27		Mohan	OK
36	P4680-01	BP-F26	SAM	11/04/24 15:29		Mohan	OK
37	P4680-05	BP-F25	SAM	11/04/24 15:31		Mohan	OK
38	P4663-02	TENNANT-PAD-2-3-S	SAM	11/04/24 15:34		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133275

Review By	mohan	Review On	11/6/2024 1:07:46 AM
Supervise By	jaswal	Supervise On	11/6/2024 1:08:59 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83017,MP83018,MP83019,MP83020,MP83021,MP83022		
ICV Standard	MP83023		
CCV Standard	MP83025		
ICSA Standard			
CRI Standard	MP83027		
LCS Standard			
Chk Standard	MP83024,MP83026,,,MP83028,MP83030		

39	P4663-02DUP	TENNANT-PAD-2-3-S	DUP	11/04/24 15:36		Mohan	OK
40	P4663-02L	TENNANT-PAD-2-3-S	SD	11/04/24 15:38		Mohan	OK
41	P4663-02A	TENNANT-PAD-2-3-S	PS	11/04/24 15:41		Mohan	OK
42	CCV25	CCV25	CCV	11/04/24 15:43		Mohan	OK
43	CCB25	CCB25	CCB	11/04/24 15:45		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133296

Review By	jaswal	Review On	11/6/2024 3:15:05 AM
Supervise By	kareem	Supervise On	11/7/2024 12:42:46 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/04/24 16:57		Kareem	OK
2	S1	S1	CAL2	11/04/24 17:02		Kareem	OK
3	S2	S2	CAL3	11/04/24 17:06		Kareem	OK
4	S3	S3	CAL4	11/04/24 17:10		Kareem	OK
5	S4	S4	CAL5	11/04/24 17:14		Kareem	OK
6	S5	S5	CAL6	11/04/24 17:18		Kareem	OK
7	ICV01	ICV01	ICV	11/04/24 17:28		Kareem	OK
8	LLICV01	LLICV01	LLICV	11/04/24 17:37		Kareem	OK
9	ICB01	ICB01	ICB	11/04/24 17:47		Kareem	OK
10	CRI01	CRI01	CRDL	11/04/24 17:52		Kareem	OK
11	ICSA01	ICSA01	ICSA	11/04/24 18:03		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	11/04/24 18:07		Kareem	OK
13	CCV01	CCV01	CCV	11/04/24 18:11		Kareem	OK
14	CCB01	CCB01	CCB	11/04/24 18:16		Kareem	OK
15	P4640-01	MH-3	SAM	11/04/24 18:20		Kareem	OK
16	P4643-01	BP-F9-ADDITIONAL	SAM	11/04/24 18:25		Kareem	OK
17	P4643-05	BP-F8	SAM	11/04/24 18:29		Kareem	OK
18	P4643-09	TP-9	SAM	11/04/24 18:33		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133296

Review By	jaswal	Review On	11/6/2024 3:15:05 AM
Supervise By	kareem	Supervise On	11/7/2024 12:42:46 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

19	P4645-01	Z-02-WC	SAM	11/04/24 18:37		Kareem	OK
20	P4645-01DUP	Z-02-WCDUP	DUP	11/04/24 18:42		Kareem	OK
21	P4645-01L	Z-02-WCL	SD	11/04/24 18:46		Kareem	OK
22	P4645-01MS	Z-02-WCMS	MS	11/04/24 18:50	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
23	P4645-01MSD	Z-02-WCMSD	MSD	11/04/24 18:54	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
24	P4645-01A	Z-02-WCA	PS	11/04/24 18:58	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
25	CCV02	CCV02	CCV	11/04/24 19:02		Kareem	OK
26	CCB02	CCB02	CCB	11/04/24 19:07		Kareem	OK
27	PB164563BL	PB164563BL	MB	11/04/24 19:11	NOT USE	Kareem	Not Ok
28	PB164563BS	PB164563BS	LCS	11/04/24 19:15	NOT USE	Kareem	Not Ok
29	P4615-03	B-131-3-SB01	SAM	11/04/24 19:19		Kareem	OK
30	P4615-02	B-131-2-SB01	SAM	11/04/24 19:23		Kareem	OK
31	P4397-04	WB-301-SW	SAM	11/04/24 19:28	Na high	Kareem	Dilution
32	P4397-04DUP	WB-301-SWDUP	DUP	11/04/24 19:32	Na high	Kareem	Dilution
33	P4397-04L	WB-301-SWL	SD	11/04/24 19:37	Na high	Kareem	Dilution
34	P4397-04MS	WB-301-SWMS	MS	11/04/24 19:41	Na high	Kareem	Dilution
35	CCV03	CCV03	CCV	11/04/24 19:46		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133296

Review By	jaswal	Review On	11/6/2024 3:15:05 AM
Supervise By	kareem	Supervise On	11/7/2024 12:42:46 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

36	CCB03	CCB03	CCB	11/04/24 19:58		Kareem	OK
37	P4397-04MSD	WB-301-SWMSD	MSD	11/04/24 20:04	Na high	Kareem	Dilution
38	P4397-04A	WB-301-SWA	PS	11/04/24 20:08	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
39	P4663-02	TENNANT-PAD-2-3-S	SAM	11/04/24 20:12		Kareem	OK
40	P4663-04	RES-BUILDING-PAD-	SAM	11/04/24 20:17		Kareem	OK
41	P4663-06	RES-BUILDING-PAD-	SAM	11/04/24 20:21		Kareem	OK
42	P4663-08	RES-BUILDING-PAD-	SAM	11/04/24 20:25		Kareem	OK
43	LR1	LR1	HIGH STD	11/04/24 20:33		Kareem	OK
44	LR2	LR2	HIGH STD	11/04/24 20:38		Kareem	OK
45	PB164522TB	PB164522TB	MB	11/04/24 20:44		Kareem	OK
46	PB164564BL	PB164564BL	MB	11/04/24 20:48		Kareem	OK
47	CCV04	CCV04	CCV	11/04/24 21:07		Kareem	OK
48	CCB04	CCB04	CCB	11/04/24 21:11		Kareem	OK
49	PB164564BS	PB164564BS	LCS	11/04/24 21:16	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
50	P4397-04DL	WB-301-SWDL	SAM	11/04/24 21:21	10X for Na	Kareem	Confirms
51	P4397-04DUPDL	WB-301-SWDUPDL	DUP	11/04/24 21:25	10X for Na	Kareem	Confirms
52	P4397-04LDL	WB-301-SWLDL	SD	11/04/24 21:29	10X for Na	Kareem	Confirms
53	P4397-04MSDL	WB-301-SWMSDL	MS	11/04/24 21:34	10X for Na	Kareem	Confirms

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133296

Review By	jaswal	Review On	11/6/2024 3:15:05 AM
Supervise By	kareem	Supervise On	11/7/2024 12:42:46 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

54	P4397-04MSDDL	WB-301-SWMSDDL	MSD	11/04/24 21:38	10X for Na	Kareem	Confirms
55	P4397-04ADL	WB-301-SWADL	PS	11/04/24 21:42	Not used	Kareem	Not Ok
56	CCV05	CCV05	CCV	11/04/24 21:47		Kareem	OK
57	CCB05	CCB05	CCB	11/04/24 21:51		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133323

Review By	kareem	Review On	11/7/2024 1:15:13 PM
Supervise By	mohan	Supervise On	11/11/2024 12:01:52 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83078 MP83079 MP73080 MP83081 MP83082 MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086 MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091 MP83092		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/06/24 17:09		Kareem	OK
2	S1	S1	CAL2	11/06/24 17:13		Kareem	OK
3	S2	S2	CAL3	11/06/24 17:17		Kareem	OK
4	S3	S3	CAL4	11/06/24 17:22		Kareem	OK
5	S4	S4	CAL5	11/06/24 17:26		Kareem	OK
6	S5	S5	CAL6	11/06/24 17:30		Kareem	OK
7	ICV01	ICV01	ICV	11/06/24 17:34		Kareem	OK
8	LLICV01	LLICV01	LLICV	11/06/24 17:38		Kareem	OK
9	ICB01	ICB01	ICB	11/06/24 17:43		Kareem	OK
10	CRI01	CRI01	CRDL	11/06/24 18:04		Kareem	OK
11	ICSA01	ICSA01	ICSA	11/06/24 18:08		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	11/06/24 18:13		Kareem	OK
13	CCV01	CCV01	CCV	11/06/24 18:23		Kareem	OK
14	CCB01	CCB01	CCB	11/06/24 18:27		Kareem	OK
15	P4709-01	HR-02-110424	SAM	11/06/24 18:31		Kareem	OK
16	P4709-03	HR-03-110424	SAM	11/06/24 18:35		Kareem	OK
17	CCV02	CCV02	CCV	11/06/24 18:39		Kareem	OK
18	CCB02	CCB02	CCB	11/06/24 18:44		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133323

Review By	kareem	Review On	11/7/2024 1:15:13 PM
Supervise By	mohan	Supervise On	11/11/2024 12:01:52 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078 MP83079 MP73080 MP83081 MP83082 MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086 MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091 MP83092

19	P4652-01	RW5-SP100-50-2024	SAM	11/06/24 18:49		Kareem	OK
20	P4679-01	MH-1	SAM	11/06/24 18:53		Kareem	OK
21	P4659-01	MH-2	SAM	11/06/24 18:57		Kareem	OK
22	P4652-03	RW5-SP301-303-50-2	SAM	11/06/24 19:02		Kareem	OK
23	P4652-04	RW5-SP100-90-2024	SAM	11/06/24 19:06		Kareem	OK
24	P4652-06	RW5-SP301-303-90-2	SAM	11/06/24 19:11		Kareem	OK
25	P4661-03	102324-A	SAM	11/06/24 19:15		Kareem	OK
26	P4680-01	BP-F26	SAM	11/06/24 19:19		Kareem	OK
27	P4680-05	BP-F25	SAM	11/06/24 19:24		Kareem	OK
28	P4680-05DUP	BP-F25DUP	DUP	11/06/24 19:28		Kareem	OK
29	CCV03	CCV03	CCV	11/06/24 19:32		Kareem	OK
30	CCB03	CCB03	CCB	11/06/24 19:36		Kareem	OK
31	P4680-05L	BP-F25L	SD	11/06/24 19:40		Kareem	OK
32	P4680-05MS	BP-F25MS	MS	11/06/24 19:44	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
33	P4680-05MSD	BP-F25MSD	MSD	11/06/24 19:48	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
34	P4680-05A	BP-F25A	PS	11/06/24 19:52	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
35	PB164634BL	PB164634BL	MB	11/06/24 19:56	NOT USE	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133323

Review By	kareem	Review On	11/7/2024 1:15:13 PM
Supervise By	mohan	Supervise On	11/11/2024 12:01:52 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078 MP83079 MP73080 MP83081 MP83082 MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086 MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091 MP83092

36	P4662-04DL	102524-DDL	SAM	11/06/24 20:05	Straight 5x Dilution for all elements	Kareem	OK
37	P4673-01	OUTFALL-001	SAM	11/06/24 20:09		Kareem	OK
38	P4673-02	OUTFALL-002	SAM	11/06/24 20:13		Kareem	OK
39	P4673-03	OUTFALL-003	SAM	11/06/24 20:18		Kareem	OK
40	CCV04	CCV04	CCV	11/06/24 20:22		Kareem	OK
41	CCB04	CCB04	CCB	11/06/24 20:26		Kareem	OK
42	P4673-04	OUTFALL-004	SAM	11/06/24 20:31		Kareem	OK
43	P4673-05	OUTFALL-005	SAM	11/06/24 20:35		Kareem	OK
44	P4673-06	OUTFALL-006	SAM	11/06/24 20:39		Kareem	OK
45	P4673-07	OUTFALL-007	SAM	11/06/24 20:44		Kareem	OK
46	P4673-07DUP	OUTFALL-007DUP	DUP	11/06/24 20:48		Kareem	OK
47	P4673-07L	OUTFALL-007L	SD	11/06/24 20:52		Kareem	OK
48	P4673-07MS	OUTFALL-007MS	MS	11/06/24 20:57	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
49	P4673-07MSD	OUTFALL-007MSD	MSD	11/06/24 21:01	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
50	P4673-07A	OUTFALL-007A	PS	11/06/24 21:05	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
51	P4674-01	OUTFALL-001	SAM	11/06/24 21:09		Kareem	OK
52	CCV05	CCV05	CCV	11/06/24 21:13		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133323

Review By	kareem	Review On	11/7/2024 1:15:13 PM
Supervise By	mohan	Supervise On	11/11/2024 12:01:52 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83078 MP83079 MP73080 MP83081 MP83082 MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086 MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091 MP83092		

53	CCB05	CCB05	CCB	11/06/24 21:17		Kareem	OK
54	P4684-03	MECHANIC-ST-OIL-V	SAM	11/06/24 21:22	Fe high & Ag (OverSaturated)	Kareem	Dilution
55	PB164645BL	PB164645BL	MB	11/06/24 21:26	NOT USE	Kareem	Not Ok
56	PB164645BS	PB164645BS	LCS	11/06/24 21:30	NOT USE - fail for Na	Kareem	Not Ok
57	P4658-01	B-131-1-SB02	SAM	11/06/24 21:34		Kareem	OK
58	P4658-02	B-131-2-SB02	SAM	11/06/24 21:39		Kareem	OK
59	P4658-03	B-131-3-SB02	SAM	11/06/24 21:43		Kareem	OK
60	P4364-15	SB-07-0.167-0.667	SAM	11/06/24 21:47	NOT USE	Kareem	Not Ok
61	P4364-16	SB-08-0.167-0.667	SAM	11/06/24 21:52	NOT USE	Kareem	Not Ok
62	P4364-14	SB-06-0.167-0.667	SAM	11/06/24 21:56		Kareem	OK
63	P4364-14DUP	SB-06-0.167-0.667DU	DUP	11/06/24 22:01		Kareem	OK
64	CCV06	CCV06	CCV	11/06/24 22:11		Kareem	OK
65	CCB06	CCB06	CCB	11/06/24 22:15		Kareem	OK
66	P4364-14L	SB-06-0.167-0.667L	SD	11/06/24 22:19		Kareem	OK
67	P4364-14MS	SB-06-0.167-0.667MS	MS	11/06/24 22:23	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
68	P4364-14MSD	SB-06-0.167-0.667MS	MSD	11/06/24 22:28	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
69	P4364-14A	SB-06-0.167-0.667A	PS	11/06/24 22:33	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133323

Review By	kareem	Review On	11/7/2024 1:15:13 PM
Supervise By	mohan	Supervise On	11/11/2024 12:01:52 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83078 MP83079 MP73080 MP83081 MP83082 MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086 MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091 MP83092		

70	P4667-01	BP-F-6	SAM	11/06/24 22:38		Kareem	OK
71	P4667-05	BP-F-5	SAM	11/06/24 22:42		Kareem	OK
72	P4667-09	TP-10	SAM	11/06/24 22:46		Kareem	OK
73	P4667-13	BP-F-7	SAM	11/06/24 22:50		Kareem	OK
74	P4675-01	COMP-1	SAM	11/06/24 22:54		Kareem	OK
75	P4675-02	COMP-2	SAM	11/06/24 22:59		Kareem	OK
76	CCV07	CCV07	CCV	11/06/24 23:03		Kareem	OK
77	CCB07	CCB07	CCB	11/06/24 23:07		Kareem	OK
78	P4675-03	COMP-3	SAM	11/06/24 23:12		Kareem	OK
79	P4675-04	COMP-4	SAM	11/06/24 23:16		Kareem	OK
80	P4675-05	COMP-5	SAM	11/06/24 23:20		Kareem	OK
81	P4675-06	COMP-6	SAM	11/06/24 23:25		Kareem	OK
82	P4685-01	OK-01-11012024	SAM	11/06/24 23:29		Kareem	OK
83	P4685-01DUP	OK-01-11012024DUP	DUP	11/06/24 23:33		Kareem	OK
84	P4685-01L	OK-01-11012024L	SD	11/06/24 23:37		Kareem	OK
85	P4685-01MS	OK-01-11012024MS	MS	11/06/24 23:41	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
86	P4685-01MSD	OK-01-11012024MSD	MSD	11/06/24 23:45	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133323

Review By	kareem	Review On	11/7/2024 1:15:13 PM
Supervise By	mohan	Supervise On	11/11/2024 12:01:52 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078 MP83079 MP73080 MP83081 MP83082 MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086 MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091 MP83092

87	P4685-01A	OK-01-11012024A	PS	11/06/24 23:49	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
88	CCV08	CCV08	CCV	11/06/24 23:59		Kareem	OK
89	CCB08	CCB08	CCB	11/07/24 00:03		Kareem	OK
90	P4397-04	WB-301-SW	SAM	11/07/24 00:07	Na (Over Saturated)	Kareem	Dilution
91	P4397-04DUP	WB-301-SWDUP	DUP	11/07/24 00:12	Na (Over Saturated)	Kareem	Dilution
92	P4397-04L	WB-301-SWL	SD	11/07/24 00:16		Kareem	OK
93	P4397-04MS	WB-301-SWMS	MS	11/07/24 00:21	Na (Over Saturated)	Kareem	Dilution
94	P4397-04MSD	WB-301-SWMSD	MSD	11/07/24 00:25	Na (Over Saturated)	Kareem	Dilution
95	P4397-04A	WB-301-SWA	PS	11/07/24 00:29	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
96	PB164647BS	PB164647BS	LCS	11/07/24 00:38	NOT USE	Kareem	Not Ok
97	LR1	LR1	HIGH STD	11/07/24 00:43		Kareem	OK
98	LR2	LR2	HIGH STD	11/07/24 00:48		Kareem	OK
99	CCV09	CCV09	CCV	11/07/24 00:53		Kareem	OK
100	CCB09	CCB09	CCB	11/07/24 00:57		Kareem	OK
101	P4397-04DL	WB-301-SWDL	SAM	11/07/24 01:01	5x for Na	Kareem	Confirms
102	P4397-04DUPDL	WB-301-SWDUPDL	DUP	11/07/24 01:06	5x for Na	Kareem	Confirms
103	P4397-04LDL	WB-301-SWLDL	SD	11/07/24 01:10	Not Required	Kareem	Not Ok
104	P4397-04MSDL	WB-301-SWMSDL	MS	11/07/24 01:15	5x for Na	Kareem	Confirms

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133323

Review By	kareem	Review On	11/7/2024 1:15:13 PM
Supervise By	mohan	Supervise On	11/11/2024 12:01:52 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83078 MP83079 MP73080 MP83081 MP83082 MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086 MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091 MP83092		

105	P4397-04MSDDL	WB-301-SWMSDDL	MSD	11/07/24 01:19	5x for Na	Kareem	Confirms
106	P4397-04ADL	WB-301-SWADL	PS	11/07/24 01:24	Not Required	Kareem	Not Ok
107	P4684-03DL	MECHANIC-ST-OIL-V	SAM	11/07/24 01:28	5x for Fe,Ag	Kareem	Confirms
108	P4364-16DL	SB-08-0.167-0.667DL	SAM	11/07/24 01:33	NOT USE	Kareem	Not Ok
109	PB164563BL	PB164563BL	MB	11/07/24 01:37	NOT USE	Kareem	Not Ok
110	PB164563BS	PB164563BS	LCS	11/07/24 01:41	NOT USE - fail for K,Na	Kareem	Not Ok
111	CCV10	CCV10	CCV	11/07/24 01:49		Kareem	OK
112	CCB10	CCB10	CCB	11/07/24 01:53		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	jaswal	Supervise On	11/9/2024 12:08:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/07/24 13:23		Kareem	OK
2	S1	S1	CAL2	11/07/24 13:28		Kareem	OK
3	S2	S2	CAL3	11/07/24 13:32		Kareem	OK
4	S3	S3	CAL4	11/07/24 13:36		Kareem	OK
5	S4	S4	CAL5	11/07/24 13:40		Kareem	OK
6	S5	S5	CAL6	11/07/24 13:45		Kareem	OK
7	ICV01	ICV01	ICV	11/07/24 13:49		Kareem	OK
8	LLICV01	LLICV01	LLICV	11/07/24 14:23		Kareem	OK
9	ICB01	ICB01	ICB	11/07/24 14:38		Kareem	OK
10	CRI01	CRI01	CRDL	11/07/24 14:42		Kareem	OK
11	ICSA01	ICSA01	ICSA	11/07/24 14:47		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	11/07/24 14:51		Kareem	OK
13	CCV01	CCV01	CCV	11/07/24 15:01		Kareem	OK
14	CCB01	CCB01	CCB	11/07/24 15:05		Kareem	OK
15	P4645-04	Z-02-WC	SAM	11/07/24 15:09		Kareem	OK
16	P4659-04	MH-2	SAM	11/07/24 15:14		Kareem	OK
17	P4660-03	WC-TA2-01-C	SAM	11/07/24 15:18		Kareem	OK
18	P4660-07	WC-WOOD-01-C	SAM	11/07/24 15:22		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	jaswal	Supervise On	11/9/2024 12:08:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

19	P4660-11	WC-CONCRETE-01-C	SAM	11/07/24 15:27		Kareem	OK
20	P4667-04	BP-F-6	SAM	11/07/24 15:31		Kareem	OK
21	P4667-08	BP-F-5	SAM	11/07/24 15:36		Kareem	OK
22	P4667-12	TP-10	SAM	11/07/24 15:40		Kareem	OK
23	P4667-16	BP-F-7	SAM	11/07/24 15:45		Kareem	OK
24	P4679-04	MH-1	SAM	11/07/24 15:49		Kareem	OK
25	CCV02	CCV02	CCV	11/07/24 15:53		Kareem	OK
26	CCB02	CCB02	CCB	11/07/24 15:58		Kareem	OK
27	P4680-04	BP-F26	SAM	11/07/24 16:02		Kareem	OK
28	P4680-08	BP-F25	SAM	11/07/24 16:06		Kareem	OK
29	P4684-01	MECHANIC-ST-SWEI	SAM	11/07/24 16:11		Kareem	OK
30	P4684-01DUP	MECHANIC-ST-SWEI	DUP	11/07/24 16:15		Kareem	OK
31	P4684-01L	MECHANIC-ST-SWEI	SD	11/07/24 16:20		Kareem	OK
32	P4684-01MS	MECHANIC-ST-SWEI	MS	11/07/24 16:24		Kareem	OK
33	P4684-01MSD	MECHANIC-ST-SWEI	MSD	11/07/24 16:28		Kareem	OK
34	P4684-01A	MECHANIC-ST-SWEI	PS	11/07/24 16:32		Kareem	OK
35	PB164560TB	PB164560TB	MB	11/07/24 16:36		Kareem	OK
36	PB164665BL	PB164665BL	MB	11/07/24 16:41		Kareem	OK
37	CCV03	CCV03	CCV	11/07/24 16:45		Kareem	OK
38	CCB03	CCB03	CCB	11/07/24 16:49		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	jaswal	Supervise On	11/9/2024 12:08:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

39	PB164665BS	PB164665BS	LCS	11/07/24 16:54		Kareem	OK
40	P4693-04	BP-G5-WC	SAM	11/07/24 16:58		Kareem	OK
41	P4693-08	BP-G4-WC	SAM	11/07/24 17:02		Kareem	OK
42	P4694-04	Z-03A	SAM	11/07/24 17:06		Kareem	OK
43	P4694-08	Z-04	SAM	11/07/24 17:11		Kareem	OK
44	P4695-04	Z-01	SAM	11/07/24 17:15		Kareem	OK
45	P4700-04	MH-8	SAM	11/07/24 17:20		Kareem	OK
46	P4701-01	BP-F3	SAM	11/07/24 17:24		Kareem	OK
47	P4701-08	BP-F4	SAM	11/07/24 17:29		Kareem	OK
48	P4711-05	CF-613-COMP-16	SAM	11/07/24 17:33		Kareem	OK
49	CCV04	CCV04	CCV	11/07/24 17:56		Kareem	OK
50	CCB04	CCB04	CCB	11/07/24 18:04		Kareem	OK
51	P4711-10	CF-613-COMP-17	SAM	11/07/24 18:08		Kareem	OK
52	P4702-01	TOTE-1	SAM	11/07/24 18:13		Kareem	OK
53	P4662-06DL2	102524-DDL2	SAM	11/07/24 18:17	NOT USE	Kareem	Not Ok
54	P4662-06DL	102524-DDL	SAM	11/07/24 18:22	NOT USE	Kareem	Not Ok
55	P4662-06	102524-D	SAM	11/07/24 18:28		Kareem	OK
56	CCV05	CCV05	CCV	11/07/24 18:32		Kareem	OK
57	CCB05	CCB05	CCB	11/07/24 18:37		Kareem	OK
58	P4702-01DUP	TOTE-1DUP	DUP	11/07/24 18:43		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	jaswal	Supervise On	11/9/2024 12:08:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

59	P4702-01L	TOTE-1L	SD	11/07/24 18:47		Kareem	OK
60	P4706-01	TR-04-110424	SAM	11/07/24 18:51		Kareem	OK
61	P4708-01	OR-02-110424	SAM	11/07/24 18:56		Kareem	OK
62	P4711-01	CF-613-COMP-16	SAM	11/07/24 19:00		Kareem	OK
63	P4711-06	CF-613-COMP-17	SAM	11/07/24 19:04		Kareem	OK
64	CCV06	CCV06	CCV	11/07/24 19:08		Kareem	OK
65	CCB06	CCB06	CCB	11/07/24 19:12		Kareem	OK
66	P4702-01MS	TOTE-1MS	MS	11/07/24 19:16		Kareem	OK
67	P4702-01MSD	TOTE-1MSD	MSD	11/07/24 19:20		Kareem	OK
68	P4702-01A	TOTE-1A	PS	11/07/24 19:24		Kareem	OK
69	PB164662TB	PB164662TB	MB	11/07/24 19:28		Kareem	OK
70	PB164685TB	PB164685TB	MB	11/07/24 19:32		Kareem	OK
71	PB164712BL	PB164712BL	MB	11/07/24 19:37		Kareem	OK
72	PB164712BS	PB164712BS	LCS	11/07/24 19:41		Kareem	OK
73	P4720-01	JC-701-COMP-01	SAM	11/07/24 19:45		Kareem	OK
74	P4720-01DUP	JC-701-COMP-01DUP	DUP	11/07/24 19:50		Kareem	OK
75	P4720-01L	JC-701-COMP-01L	SD	11/07/24 19:54		Kareem	OK
76	CCV07	CCV07	CCV	11/07/24 19:58		Kareem	OK
77	CCB07	CCB07	CCB	11/07/24 20:02		Kareem	OK
78	P4617-04	CONCRETE-PILE	SAM	11/07/24 20:08		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	jaswal	Supervise On	11/9/2024 12:08:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

79	P4617-04DUP	CONCRETE-PILEDU	DUP	11/07/24 20:12		Kareem	OK
80	P4617-04L	CONCRETE-PILEL	SD	11/07/24 20:17		Kareem	OK
81	P4617-04MS	CONCRETE-PILEMS	MS	11/07/24 20:21		Kareem	OK
82	P4617-04MSD	CONCRETE-PILEMS	MSD	11/07/24 20:25		Kareem	OK
83	P4549-04	TT-069-IDWGW-2024	SAM	11/07/24 20:34		Kareem	OK
84	P4617-04A	CONCRETE-PILEA	PS	11/07/24 20:41		Kareem	OK
85	P4549-04DUP	TT-069-IDWGW-2024	DUP	11/07/24 20:45		Kareem	OK
86	P4549-04L	TT-069-IDWGW-2024	SD	11/07/24 20:49		Kareem	OK
87	P4549-04MS	TT-069-IDWGW-2024	MS	11/07/24 20:54		Kareem	OK
88	CCV08	CCV08	CCV	11/07/24 20:58		Kareem	OK
89	CCB08	CCB08	CCB	11/07/24 21:02		Kareem	OK
90	P4549-04MSD	TT-069-IDWGW-2024	MSD	11/07/24 21:08		Kareem	OK
91	P4549-04A	TT-069-IDWGW-2024	PS	11/07/24 21:12		Kareem	OK
92	P4720-01MS	JC-701-COMP-01MS	MS	11/07/24 21:16		Kareem	OK
93	P4720-01MSD	JC-701-COMP-01MS	MSD	11/07/24 21:20		Kareem	OK
94	P4720-01A	JC-701-COMP-01A	PS	11/07/24 21:24		Kareem	OK
95	P4722-03	WC-1(0-6)	SAM	11/07/24 21:29	Fe high	Kareem	Dilution
96	P4722-08	WC-2(0-6)	SAM	11/07/24 21:33		Kareem	OK
97	P4722-13	WC-3(0-6)	SAM	11/07/24 21:38		Kareem	OK
98	PB164723BL	PB164723BL	MB	11/07/24 21:42		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	jaswal	Supervise On	11/9/2024 12:08:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

99	PB164723BS	PB164723BS	LCS	11/07/24 21:47		Kareem	OK
100	CCV09	CCV09	CCV	11/07/24 21:51		Kareem	OK
101	CCB09	CCB09	CCB	11/07/24 21:55		Kareem	OK
102	P4693-01	BP-G5-WC	SAM	11/07/24 22:09		Kareem	OK
103	P4693-05	BP-G4-WC	SAM	11/07/24 22:13		Kareem	OK
104	P4694-01	Z-03A	SAM	11/07/24 22:17		Kareem	OK
105	P4694-05	Z-04	SAM	11/07/24 22:22		Kareem	OK
106	P4695-01	Z-01	SAM	11/07/24 22:26		Kareem	OK
107	P4697-01	TP-1	SAM	11/07/24 22:30		Kareem	OK
108	P4699-01	MIXED-DEMO	SAM	11/07/24 22:34		Kareem	OK
109	P4699-01DUP	MIXED-DEMODUP	DUP	11/07/24 22:39		Kareem	OK
110	P4699-01L	MIXED-DEMOL	SD	11/07/24 22:43		Kareem	OK
111	P4699-01MS	MIXED-DEMOMS	MS	11/07/24 22:47		Kareem	OK
112	CCV10	CCV10	CCV	11/07/24 22:52		Kareem	OK
113	CCB10	CCB10	CCB	11/07/24 22:56		Kareem	OK
114	P4699-01MSD	MIXED-DEMOMSD	MSD	11/07/24 23:00		Kareem	OK
115	P4699-01A	MIXED-DEMOA	PS	11/07/24 23:04		Kareem	OK
116	P4700-01	MH-8	SAM	11/07/24 23:09		Kareem	OK
117	P4701-01RE	BP-F3RE	SAM	11/07/24 23:13	NOT USE	Kareem	Not Ok
118	P4701-05	BP-F4	SAM	11/07/24 23:17		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	jaswal	Supervise On	11/9/2024 12:08:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

119	P4707-01	HD-02-110424	SAM	11/07/24 23:22		Kareem	OK
120	P4718-01	WB-307-SB01	SAM	11/07/24 23:26		Kareem	OK
121	P4718-02	WB-307-SB02	SAM	11/07/24 23:30		Kareem	OK
122	P4719-01	BAYAVE-STOCKPILE	SAM	11/07/24 23:34		Kareem	OK
123	PB164708BL	PB164708BL	MB	11/07/24 23:38		Kareem	OK
124	CCV11	CCV11	CCV	11/07/24 23:43		Kareem	OK
125	CCB11	CCB11	CCB	11/07/24 23:47		Kareem	OK
126	PB164708BS	PB164708BS	LCS	11/07/24 23:51		Kareem	OK
127	PB164563BL	PB164563BL	MB	11/07/24 23:55		Kareem	OK
128	PB164563BS	PB164563BS	LCS	11/08/24 00:05		Kareem	OK
129	PB164634BL	PB164634BL	MB	11/08/24 00:08		Kareem	OK
130	PB164634BS	PB164634BS	LCS	11/08/24 00:13		Kareem	OK
131	PB164647BL	PB164647BL	MB	11/08/24 00:17		Kareem	OK
132	PB164647BS	PB164647BS	LCS	11/08/24 00:21		Kareem	OK
133	P4722-03DL	WC-1(0-6)DL	SAM	11/08/24 00:26	5x for Fe	Kareem	Confirms
134	LR1	LR1	HIGH STD	11/08/24 00:30		Kareem	OK
135	LR2	LR2	HIGH STD	11/08/24 00:35		Kareem	OK
136	CCV12	CCV12	CCV	11/08/24 00:39		Kareem	OK
137	CCB12	CCB12	CCB	11/08/24 00:43		Kareem	OK
138	P4701-01DL	BP-F3DL	SAM	11/08/24 00:48	NOT USE	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133344

Review By	mohan	Review On	11/9/2024 12:07:42 AM
Supervise By	jaswal	Supervise On	11/9/2024 12:08:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

139	CCV13	CCV13	CCV	11/08/24 00:57		Kareem	OK
140	CCB13	CCB13	CCB	11/08/24 01:01		Kareem	OK