

Continuing Calibration Report for SequenceID : FE110124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE050990.D

Aliphatic C9-C12	7907796.000	60.000	3.143	6.761	131796.600	142899.930	7.770
Aliphatic C12-C16	5406989.000	40.000	6.762	10.193	135174.725	145118.273	6.852
Aliphatic C16-C21	8107839.000	60.000	10.194	13.552	135130.650	144401.245	6.420
Aliphatic C21-C28	10983169.000	80.000	13.553	17.208	137289.613	142154.874	3.423
Aliphatic C28-C40	16792590.000	120.000	17.209	22.051	139938.250	138303.281	-1.182
Aliphatic EPH	49198383.000	360.000	3.143	22.051	136662.175	141698.847	3.554