

Continuing Calibration Report for SequenceID : FE110524AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FE051023.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 7665518.000  | 60.000  | 3.140  | 6.760  | 127758.633 | 142899.930 | 10.596 |
| Aliphatic C12-C16 | 5082769.000  | 40.000  | 6.761  | 10.191 | 127069.225 | 145118.273 | 12.437 |
| Aliphatic C16-C21 | 7446348.000  | 60.000  | 10.192 | 13.549 | 124105.800 | 144401.245 | 14.055 |
| Aliphatic C21-C28 | 9517574.000  | 80.000  | 13.550 | 17.205 | 118969.675 | 142154.874 | 16.310 |
| Aliphatic C28-C40 | 13061518.000 | 120.000 | 17.206 | 22.048 | 108845.983 | 138303.281 | 21.299 |
| Aliphatic EPH     | 42773727.000 | 360.000 | 3.140  | 22.048 | 118815.908 | 141698.847 | 16.149 |