

### CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

Continuing Calib Date: 11/12/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 06:59 Initial Calibration Time(s): 15:13 16:58

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| DICAMBA          | 7.43       | 7.44      | 7.34      | 7.54  | 0.01       |
| 2,4-DCAA         | 7.24       | 7.25      | 7.15      | 7.35  | 0.01       |
| DICHLORPROP      | 8.14       | 8.15      | 8.05      | 8.25  | 0.01       |
| 2,4-D            | 8.37       | 8.38      | 8.28      | 8.48  | 0.01       |
| 2,4,5-TP(Silvex) | 9.25       | 9.26      | 9.16      | 9.36  | 0.01       |
| 2,4,5-T          | 9.54       | 9.55      | 9.45      | 9.65  | 0.01       |
| 2,4-DB           | 10.12      | 10.13     | 10.03     | 10.23 | 0.01       |
| Dinoseb          | 11.33      | 11.35     | 11.25     | 11.45 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

Continuing Calib Date: 11/12/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 06:59 Initial Calibration Time(s): 15:13 16:58

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| DICAMBA          | 7.95       | 7.96      | 7.86      | 8.06  | 0.01       |
| 2,4-DCAA         | 7.75       | 7.76      | 7.66      | 7.86  | 0.01       |
| DICHLORPROP      | 8.67       | 8.67      | 8.57      | 8.77  | 0.01       |
| 2,4-D            | 9.00       | 9.01      | 8.91      | 9.11  | 0.02       |
| 2,4,5-TP(Silvex) | 9.90       | 9.91      | 9.81      | 10.01 | 0.01       |
| 2,4,5-T          | 10.32      | 10.33     | 10.23     | 10.43 | 0.01       |
| 2,4-DB           | 10.89      | 10.90     | 10.80     | 11.00 | 0.01       |
| Dinoseb          | 11.27      | 11.28     | 11.18     | 11.38 | 0.01       |