

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/05/2024 11/05/2024

Client Sample No.: CCAL01 Date Analyzed: 11/05/2024

Lab Sample No.: HSTDCCC750 Data File : PS028225.D Time Analyzed: 20:36

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.572	9.473	9.673	760.150	712.500	6.7
2,4,5-TP(Silvex)	9.280	9.181	9.381	761.980	712.500	6.9
2,4-D	8.397	8.298	8.498	729.750	705.000	3.5
2,4-DB	10.150	10.051	10.251	744.440	712.500	4.5
2,4-DCAA	7.265	7.166	7.366	757.410	750.000	1.0
DICAMBA	7.455	7.355	7.555	722.390	705.000	2.5
DICHLORPROP	8.166	8.066	8.266	724.670	705.000	2.8
Dinoseb	11.367	11.268	11.468	760.440	705.000	7.9

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/05/2024 11/05/2024

Client Sample No.: CCAL01 Date Analyzed: 11/05/2024

Lab Sample No.: HSTDCCC750 Data File : PS028225.D Time Analyzed: 20:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.348	10.248	10.448	781.700	712.500	9.7
2,4,5-TP(Silvex)	9.926	9.826	10.026	788.840	712.500	10.7
2,4-D	9.019	8.919	9.119	776.830	705.000	10.2
2,4-DB	10.916	10.815	11.015	759.470	712.500	6.6
2,4-DCAA	7.768	7.667	7.867	823.610	750.000	9.8
DICAMBA	7.970	7.869	8.069	792.350	705.000	12.4
DICHLORPROP	8.687	8.587	8.787	781.350	705.000	10.8
Dinoseb	11.296	11.196	11.396	772.370	705.000	9.6