

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL03 Date Analyzed: 11/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS028406.D Time Analyzed: 16:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.547	9.452	9.652	727.150	712.500	2.1
2,4,5-TP(Silvex)	9.255	9.160	9.360	727.880	712.500	2.2
2,4-D	8.375	8.279	8.479	706.140	705.000	0.2
2,4-DB	10.124	10.029	10.229	698.650	712.500	-1.9
2,4-DCAA	7.246	7.150	7.350	749.600	750.000	-0.1
DICAMBA	7.434	7.338	7.538	723.500	705.000	2.6
DICHLORPROP	8.144	8.048	8.248	710.450	705.000	0.8
Dinoseb	11.339	11.245	11.445	715.780	705.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL03 Date Analyzed: 11/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS028406.D Time Analyzed: 16:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.328	10.233	10.433	676.610	712.500	-5.0
2,4,5-TP(Silvex)	9.908	9.811	10.011	676.600	712.500	-5.0
2,4-D	9.000	8.905	9.105	665.620	705.000	-5.6
2,4-DB	10.895	10.800	11.000	657.840	712.500	-7.7
2,4-DCAA	7.752	7.655	7.855	697.880	750.000	-6.9
DICAMBA	7.953	7.856	8.056	674.960	705.000	-4.3
DICHLORPROP	8.669	8.573	8.773	656.080	705.000	-6.9
Dinoseb	11.276	11.180	11.380	665.940	705.000	-5.5