

### CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL04 Date Analyzed: 11/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS028418.D Time Analyzed: 21:22

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|-----------|--------|--------------------|-------------------|------|
|                  |        | FROM      | TO     |                    |                   |      |
| 2,4,5-T          | 9.547  | 9.452     | 9.652  | 736.420            | 712.500           | 3.4  |
| 2,4,5-TP(Silvex) | 9.254  | 9.160     | 9.360  | 736.480            | 712.500           | 3.4  |
| 2,4-D            | 8.374  | 8.279     | 8.479  | 714.440            | 705.000           | 1.3  |
| 2,4-DB           | 10.123 | 10.029    | 10.229 | 702.180            | 712.500           | -1.4 |
| 2,4-DCAA         | 7.245  | 7.150     | 7.350  | 742.700            | 750.000           | -1.0 |
| DICAMBA          | 7.434  | 7.338     | 7.538  | 731.350            | 705.000           | 3.7  |
| DICHLORPROP      | 8.143  | 8.048     | 8.248  | 717.160            | 705.000           | 1.7  |
| Dinoseb          | 11.337 | 11.245    | 11.445 | 726.290            | 705.000           | 3.0  |

### CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL04 Date Analyzed: 11/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS028418.D Time Analyzed: 21:22

| COMPOUND         | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|----------------------|--------|--------------------|-------------------|------|
| 2,4,5-T          | 10.327 | 10.233               | 10.433 | 678.220            | 712.500           | -4.8 |
| 2,4,5-TP(Silvex) | 9.906  | 9.811                | 10.011 | 677.390            | 712.500           | -4.9 |
| 2,4-D            | 9.000  | 8.905                | 9.105  | 636.580            | 705.000           | -9.7 |
| 2,4-DB           | 10.894 | 10.800               | 11.000 | 666.670            | 712.500           | -6.4 |
| 2,4-DCAA         | 7.751  | 7.655                | 7.855  | 698.440            | 750.000           | -6.9 |
| DICAMBA          | 7.952  | 7.856                | 8.056  | 673.550            | 705.000           | -4.5 |
| DICHLORPROP      | 8.669  | 8.573                | 8.773  | 656.760            | 705.000           | -6.8 |
| Dinoseb          | 11.274 | 11.180               | 11.380 | 666.370            | 705.000           | -5.5 |