

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110524\
 Data File : PS028237.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2024 01:24
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 10:45:41 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110524.M
 Quant Title : 8080.M
 QLast Update : Wed Nov 06 10:44:06 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.265	7.767	1760.1E6	1233.1E6	976.190	788.492
Target Compounds							
1) T	Dalapon	2.656	2.711	2243.5E6	1984.0E6	704.231	716.877
2) T	3,5-DICHL...	6.432	6.717	2197.8E6	1752.2E6	784.204	747.844
3) T	4-Nitroph...	7.064	7.294	928.1E6	763.6E6	783.023	721.377
5) T	DICAMBA	7.454	7.969	6608.1E6	5431.4E6	836.617	753.026
6) T	MCPD	7.638	8.071	458.9E6	401.5E6	83.332	77.582
7) T	MCPA	7.789	8.317	634.1E6	534.4E6	78.175	74.508
8) T	DICHLORPROP	8.166	8.686	1549.3E6	1277.6E6	857.566	733.989
9) T	2,4-D	8.397	9.018	1749.4E6	1417.1E6	862.884	726.722
10) T	Pentachlo...	8.697	9.549	24027.8E6	18899.0E6	898.371	730.534
11) T	2,4,5-TP ...	9.279	9.925	9304.4E6	7727.5E6	929.584	718.683
12) T	2,4,5-T	9.572	10.347	9460.8E6	7598.7E6	925.988	710.287
13) T	2,4-DB	10.150	10.915	1540.0E6	955.5E6	933.839	689.345 #
14) T	DINOSEB	11.367	11.295	7545.1E6	5262.9E6	1004.201	701.605 #
15) T	Picloram	11.172	12.396	16421.3E6	11313.2E6	941.081	689.087 #
16) T	DCPA	11.659	12.343	12843.1E6	8431.2E6	1031.584	668.938 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110524\
Data File : PS028237.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2024 01:24
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 10:45:41 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110524.M
Quant Title : 8080.M
QLast Update : Wed Nov 06 10:44:06 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

