

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110924\
 Data File : PS028360.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Nov 2024 16:58
 Operator : AR\AJ
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 17:45:49 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110924.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 08 17:45:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.250	7.755	3767.0E6	2310.6E6	1364.492	1450.086
Target Compounds							
1) T	Dalapon	2.650	2.707	4944.0E6	3823.9E6	1385.601	1348.490
2) T	3,5-DICHL...	6.420	6.707	5107.0E6	3191.5E6	1276.301	1347.181
3) T	4-Nitroph...	7.049	7.281	2324.2E6	1319.3E6	1253.460	1245.836
5) T	DICAMBA	7.439	7.957	16040.0E6	10110.0E6	1332.146	1410.143
6) T	MCPD	7.628	8.064	1220.9E6	835.8E6	149.279	142.280
7) T	MCPA	7.779	8.312	1574.5E6	1084.0E6	140.433	135.369
8) T	DICHLORPROP	8.149	8.674	4014.9E6	2452.1E6	1295.283	1376.618
9) T	2,4-D	8.380	9.005	4510.1E6	2651.6E6	1283.392	1371.186
10) T	Pentachlo...	8.683	9.536	48275.3E6	32737.8E6	1071.782	1313.882
11) T	2,4,5-TP ...	9.261	9.912	24679.1E6	14436.4E6	1303.775	1362.929
12) T	2,4,5-T	9.553	10.334	25233.9E6	14128.2E6	1301.992	1361.617
13) T	2,4-DB	10.130	10.900	4217.5E6	1892.7E6	1364.499	1428.147
14) T	DINOSEB	11.345	11.280	21251.0E6	9772.8E6	1310.710	1361.620
15) T	Picloram	11.150	12.379	42075.2E6	20561.5E6	1339.720	1428.629
16) T	DCPA	11.638	12.326	35013.3E6	15448.5E6	1310.064	1378.083

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110924\
Data File : PS028360.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Nov 2024 16:58
Operator : AR\AJ
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 17:45:49 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110924.M
Quant Title : 8080.M
QLast Update : Fri Nov 08 17:45:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

