

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110524

Review By	Abdul	Review On	11/6/2024 10:27:27 AM
Supervise By	Ankita	Supervise On	11/6/2024 1:07:27 PM
SubDirectory	PS110524	HP Acquire Method	HP Processing Method ps110524 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469
CCC	PP23462
Internal Standard/PEM	
ICV/I.BLK	PP23469
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028216.D	05 Nov 2024 16:59		AR\AJ	Ok
2	I.BLK	I.BLK	PS028217.D	05 Nov 2024 17:23		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS028218.D	05 Nov 2024 17:47		AR\AJ	Ok,M
4	HSTDICC500	HSTDICC500	PS028219.D	05 Nov 2024 18:11		AR\AJ	Ok,M
5	HSTDICC750	HSTDICC750	PS028220.D	05 Nov 2024 18:36		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS028221.D	05 Nov 2024 18:59		AR\AJ	Ok,M
7	HSTDICC1500	HSTDICC1500	PS028222.D	05 Nov 2024 19:23		AR\AJ	Ok
8	HSTDICV750	ICVPS110524	PS028223.D	05 Nov 2024 19:48		AR\AJ	Ok
9	I.BLK	I.BLK	PS028224.D	05 Nov 2024 20:12		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS028225.D	05 Nov 2024 20:36		AR\AJ	Ok
11	P4663-02	TENNANT-PAD-2-3-ST	PS028226.D	05 Nov 2024 21:00		AR\AJ	Ok,M
12	P4663-04	RES-BUILDING-PAD-N	PS028227.D	05 Nov 2024 21:24		AR\AJ	Ok,M
13	P4663-06	RES-BUILDING-PAD-N	PS028228.D	05 Nov 2024 21:48		AR\AJ	Ok,M
14	P4663-08	RES-BUILDING-PAD-N	PS028229.D	05 Nov 2024 22:12		AR\AJ	Ok,M
15	PB164641BL	PB164641BL	PS028230.D	05 Nov 2024 22:36		AR\AJ	Ok,M
16	PB164641BS	PB164641BS	PS028231.D	05 Nov 2024 23:00		AR\AJ	Ok,M
17	P4640-01	MH-3	PS028232.D	05 Nov 2024 23:24		AR\AJ	Ok,M
18	P4643-01	BP-F9-ADDITIONAL	PS028233.D	05 Nov 2024 23:48		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110524

Review By	Abdul	Review On	11/6/2024 10:27:27 AM
Supervise By	Ankita	Supervise On	11/6/2024 1:07:27 PM
SubDirectory	PS110524	HP Acquire Method	HP Processing Method ps110524 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC	PP23462		
Internal Standard/PEM	PP23469		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P4643-05	BP-F8	PS028234.D	06 Nov 2024 00:12		AR\AJ	Ok,M
20	P4643-09	TP-9	PS028235.D	06 Nov 2024 00:36	Surrogate Fail	AR\AJ	ReRun
21	I.BLK	I.BLK	PS028236.D	06 Nov 2024 01:00		AR\AJ	Ok,M
22	HSTDCCC750	HSTDCCC750	PS028237.D	06 Nov 2024 01:24	CCC failling in higher side	AR\AJ	Ok
23	P4645-01	Z-02-WC	PS028238.D	06 Nov 2024 01:48		AR\AJ	Ok,M
24	P4659-01	MH-2	PS028239.D	06 Nov 2024 02:12		AR\AJ	Ok,M
25	P4602-43	CC0R8	PS028240.D	06 Nov 2024 04:36		AR\AJ	Not Ok
26	P4602-44MS	CC0R8MS	PS028241.D	06 Nov 2024 05:00	Some compound recovery fail	AR\AJ	Not Ok
27	P4602-45MSD	CC0R8MSD	PS028242.D	06 Nov 2024 05:24	Some compound recovery fail	AR\AJ	Not Ok
28	PB164519BL	PB164519BL	PS028243.D	06 Nov 2024 05:48		AR\AJ	Not Ok
29	PB164519BS	PB164519BS	PS028244.D	06 Nov 2024 06:12	Some compound recovery fail	AR\AJ	Not Ok
30	PB164393TB	PB164393TB	PS028245.D	06 Nov 2024 06:36	Surrogate Fail	AR\AJ	Not Ok
31	P4578-06	WB-305-BOT	PS028246.D	06 Nov 2024 07:00		AR\AJ	Not Ok
32	I.BLK	I.BLK	PS028247.D	06 Nov 2024 07:24		AR\AJ	Ok
33	HSTDCCC750	HSTDCCC750	PS028248.D	06 Nov 2024 07:48	CCC failling in higher side	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110924

Review By	Abdul	Review On	11/11/2024 5:07:36 PM
Supervise By	Ankita	Supervise On	11/12/2024 8:59:40 AM
SubDirectory	PS110924	HP Acquire Method	HP Processing Method ps110924 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469
CCC	PP23462
Internal Standard/PEM	
ICV/I.BLK	PP23469
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028354.D	08 Nov 2024 14:00		AR\AJ	Ok
2	I.BLK	I.BLK	PS028355.D	08 Nov 2024 14:25		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS028356.D	08 Nov 2024 15:13		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS028357.D	08 Nov 2024 15:46		AR\AJ	Ok,M
5	HSTDICC750	HSTDICC750	PS028358.D	08 Nov 2024 16:10		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS028359.D	08 Nov 2024 16:34		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS028360.D	08 Nov 2024 16:58		AR\AJ	Ok
8	HSTDICV750	ICVPS110924	PS028361.D	08 Nov 2024 17:22		AR\AJ	Ok
9	I.BLK	I.BLK	PS028362.D	08 Nov 2024 17:46		AR\AJ	Ok,M
10	HSTDCCC750	HSTDCCC750	PS028363.D	08 Nov 2024 18:10		AR\AJ	Ok
11	P4700-01	MH-8	PS028364.D	08 Nov 2024 18:34		AR\AJ	Ok,M
12	P4701-01	BP-F3	PS028365.D	08 Nov 2024 18:58		AR\AJ	Ok,M
13	P4701-05	BP-F4	PS028366.D	08 Nov 2024 19:22		AR\AJ	Ok
14	PB164754BL	PB164754BL	PS028367.D	08 Nov 2024 19:46		AR\AJ	Ok
15	PB164754BS	PB164754BS	PS028368.D	08 Nov 2024 20:10		AR\AJ	Ok,M
16	P4722-03	WC-1(0-6)	PS028369.D	08 Nov 2024 20:34	2,4-DCAA low in both column	AR\AJ	Not Ok
17	P4722-08	WC-2(0-6)	PS028370.D	08 Nov 2024 20:58		AR\AJ	Ok,M
18	P4722-13	WC-3(0-6)	PS028371.D	08 Nov 2024 21:22	2,4-DCAA Low in both column	AR\AJ	Not Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110924

Review By	Abdul	Review On	11/11/2024 5:07:36 PM
Supervise By	Ankita	Supervise On	11/12/2024 8:59:40 AM
SubDirectory	PS110924	HP Acquire Method	HP Processing Method ps110924 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC	PP23462		
Internal Standard/PEM			
ICV/I.BLK	PP23469		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P4739-01	TP-14	PS028372.D	08 Nov 2024 21:46		AR\AJ	Ok,M
20	P4739-01MS	TP-14MS	PS028373.D	08 Nov 2024 22:10	Some compound recovery fail	AR\AJ	Ok,M
21	P4739-01MSD	TP-14MSD	PS028374.D	08 Nov 2024 22:34	Some compound recovery fail	AR\AJ	Ok,M
22	P4739-05	BP-G2	PS028375.D	08 Nov 2024 22:58		AR\AJ	Ok
23	P4739-09	BP-B2	PS028376.D	08 Nov 2024 23:22		AR\AJ	Ok,M
24	P4739-13	TP-11	PS028377.D	08 Nov 2024 23:46		AR\AJ	Ok,M
25	I.BLK	I.BLK	PS028378.D	09 Nov 2024 00:10		AR\AJ	Ok
26	HSTDCCC750	HSTDCCC750	PS028379.D	09 Nov 2024 00:34		AR\AJ	Ok
27	P4601-24RE	CC0P1RE	PS028380.D	09 Nov 2024 00:58	Closing CCC fail,	AR\AJ	Not Ok
28	P4601-25RE	CC0P3RE	PS028381.D	09 Nov 2024 01:22	Closing CCC fail,	AR\AJ	Not Ok
29	P4601-26RE	CC0P5RE	PS028382.D	09 Nov 2024 01:46	Closing CCC fail,	AR\AJ	Not Ok
30	P4602-43	CC0R8	PS028383.D	09 Nov 2024 02:10	, Closing CCC fail	AR\AJ	Not Ok
31	P4602-44MS	CC0R8MS	PS028384.D	09 Nov 2024 02:34	Some compound recovery fail, Closing CCC fail	AR\AJ	Not Ok
32	P4602-45MSD	CC0R8MSD	PS028385.D	09 Nov 2024 02:58	Some compound recovery fail, Closing CCC fail	AR\AJ	Not Ok
33	P4602-40	CC0R6	PS028386.D	09 Nov 2024 03:22	Closing CCC fail	AR\AJ	Not Ok
34	P4602-41MS	CC0R6MS	PS028387.D	09 Nov 2024 03:46	Some compound recovery fail, Closing CCC fail	AR\AJ	Not Ok
35	P4602-42MSD	CC0R6MSD	PS028388.D	09 Nov 2024 04:10	Some compound recovery fail, RPD fail, Closing CCC fail	AR\AJ	Not Ok
36	PB164393TB	PB164393TB	PS028389.D	09 Nov 2024 04:34	Closing CCC fail	AR\AJ	Not Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS110924

Review By	Abdul	Review On	11/11/2024 5:07:36 PM
Supervise By	Ankita	Supervise On	11/12/2024 8:59:40 AM
SubDirectory	PS110924	HP Acquire Method	HP Processing Method ps110924 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469
CCC Internal Standard/PEM	PP23462
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469

37	I.BLK	I.BLK	PS028390.D	09 Nov 2024 04:58	Closing CCC fail	AR\AJ	Not Ok
38	HSTDCCC750	HSTDCCC750	PS028391.D	09 Nov 2024 05:22	CCC fail	AR\AJ	Not Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS111124

Review By	yogesh	Review On	11/12/2024 5:16:17 AM
Supervise By	Ankita	Supervise On	11/12/2024 9:00:51 AM
SubDirectory	PS111124	HP Acquire Method	HP Processing Method ps110924 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469
CCC	PP23462
Internal Standard/PEM	
ICV/I.BLK	PP23469
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028392.D	11 Nov 2024 10:00		AR\AJ	Ok
2	I.BLK	I.BLK	PS028393.D	11 Nov 2024 10:24		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS028394.D	11 Nov 2024 10:48		AR\AJ	Ok,M
4	P4601-29	CC0Q1	PS028395.D	11 Nov 2024 12:05		AR\AJ	Ok
5	P4601-30	CC0Q6	PS028396.D	11 Nov 2024 12:29		AR\AJ	Ok
6	P4601-31	CC0Q8	PS028397.D	11 Nov 2024 12:56		AR\AJ	Ok
7	P4601-32	CC0R3	PS028398.D	11 Nov 2024 13:20		AR\AJ	Ok
8	P4601-33	CC0R4	PS028399.D	11 Nov 2024 13:44		AR\AJ	Ok,M
9	P4601-34	CC0R5	PS028400.D	11 Nov 2024 14:08		AR\AJ	Ok,M
10	P4601-35	CC0R6	PS028401.D	11 Nov 2024 14:32		AR\AJ	Ok,M
11	P4601-36	CC0R7	PS028402.D	11 Nov 2024 14:57		AR\AJ	Ok,M
12	P4601-37MS	CC0R7MS	PS028403.D	11 Nov 2024 15:21	some comp. recovery fail	AR\AJ	Ok,M
13	P4601-38MSD	CC0R7MSD	PS028404.D	11 Nov 2024 15:45	some comp. recovery fail	AR\AJ	Ok,M
14	I.BLK	I.BLK	PS028405.D	11 Nov 2024 16:09		AR\AJ	Ok
15	HSTDCCC750	HSTDCCC750	PS028406.D	11 Nov 2024 16:33		AR\AJ	Ok,M
16	P4602-43	CC0R8	PS028407.D	11 Nov 2024 16:57		AR\AJ	Ok
17	P4602-44MS	CC0R8MS	PS028408.D	11 Nov 2024 17:21		AR\AJ	Ok,M
18	P4602-45MSD	CC0R8MSD	PS028409.D	11 Nov 2024 17:45		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS111124

Review By	yogesh	Review On	11/12/2024 5:16:17 AM
Supervise By	Ankita	Supervise On	11/12/2024 9:00:51 AM
SubDirectory	PS111124	HP Acquire Method	HP Processing Method ps110924 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC	PP23462		
Internal Standard/PEM			
ICV/I.BLK	PP23469		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P4602-40	CC0R6	PS028410.D	11 Nov 2024 18:10		AR\AJ	Ok
20	P4602-41MS	CC0R6MS	PS028411.D	11 Nov 2024 18:34		AR\AJ	Ok,M
21	P4602-42MSD	CC0R6MSD	PS028412.D	11 Nov 2024 18:58		AR\AJ	Ok,M
22	PB164393TB	PB164393TB	PS028413.D	11 Nov 2024 19:22		AR\AJ	Ok
23	PB164725BL	PB164725BL	PS028414.D	11 Nov 2024 19:46		AR\AJ	Ok
24	PB164725BS	PB164725BS	PS028415.D	11 Nov 2024 20:10		AR\AJ	Ok,M
25	P4602-19	C0P17	PS028416.D	11 Nov 2024 20:34		AR\AJ	Ok,M
26	I.BLK	I.BLK	PS028417.D	11 Nov 2024 20:58		AR\AJ	Ok
27	HSTDCCC750	HSTDCCC750	PS028418.D	11 Nov 2024 21:22		AR\AJ	Ok,M
28	P4602-21	CC0P2	PS028419.D	11 Nov 2024 21:46		AR\AJ	Ok,M
29	P4602-22	CC0P4	PS028420.D	11 Nov 2024 22:10		AR\AJ	Ok,M
30	P4602-23	CC0P6	PS028421.D	11 Nov 2024 22:34		AR\AJ	Ok
31	P4602-24	CC0P8	PS028422.D	11 Nov 2024 22:58		AR\AJ	Ok,M
32	P4602-25	CC0Q0	PS028423.D	11 Nov 2024 23:22		AR\AJ	Ok,M
33	P4602-26	CC0Q2	PS028424.D	11 Nov 2024 23:46		AR\AJ	Ok,M
34	P4602-27	CC0Q3	PS028425.D	12 Nov 2024 00:10		AR\AJ	Ok,M
35	P4602-28	CC0Q4	PS028426.D	12 Nov 2024 00:34	injection error	AR\AJ	Not Ok
36	P4602-29	CC0Q5	PS028427.D	12 Nov 2024 00:58		AR\AJ	Ok,M
37	P4602-30	CC0Q7	PS028428.D	12 Nov 2024 01:23		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS111124

Review By	yogesh	Review On	11/12/2024 5:16:17 AM
Supervise By	Ankita	Supervise On	11/12/2024 9:00:51 AM
SubDirectory	PS111124	HP Acquire Method	HP Processing Method ps110924 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC	PP23462		
Internal Standard/PEM			
ICV/I.BLK	PP23469		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	I.BLK	I.BLK	PS028429.D	12 Nov 2024 01:47		AR\AJ	Ok
39	HSTDCCC750	HSTDCCC750	PS028430.D	12 Nov 2024 02:11		AR\AJ	Ok
40	P4602-33	CC0R1	PS028431.D	12 Nov 2024 02:35		AR\AJ	Ok,M
41	P4602-34	CC0R2	PS028432.D	12 Nov 2024 02:59		AR\AJ	Ok,M
42	P4602-31	CC0Q9	PS028433.D	12 Nov 2024 03:23		AR\AJ	Ok,M
43	P4602-32	CC0R0	PS028434.D	12 Nov 2024 03:47		AR\AJ	Ok,M
44	P4602-37	CC0P0	PS028435.D	12 Nov 2024 04:11		AR\AJ	Ok,M
45	P4602-38MS	CC0P0MS	PS028436.D	12 Nov 2024 04:35	Some compound recovery fail	AR\AJ	Ok,M
46	P4602-39MSD	CC0P0MSD	PS028437.D	12 Nov 2024 04:59	Some compound recovery fail	AR\AJ	Ok,M
47	P4722-03	WC-1(0-6)	PS028438.D	12 Nov 2024 05:23		AR\AJ	Ok
48	P4722-13	WC-3(0-6)	PS028439.D	12 Nov 2024 05:47		AR\AJ	Ok
49	P4738-02	72-12018	PS028440.D	12 Nov 2024 06:11		AR\AJ	Ok
50	I.BLK	I.BLK	PS028441.D	12 Nov 2024 06:35		AR\AJ	Ok
51	HSTDCCC750	HSTDCCC750	PS028442.D	12 Nov 2024 06:59		AR\AJ	Ok,M
52	P4789-01	BP-F21	PS028443.D	12 Nov 2024 07:23		AR\AJ	Ok
53	P4792-03	OILY-STONE-DRUM	PS028444.D	12 Nov 2024 07:47		AR\AJ	Ok
54	P4738-02	72-12018	PS028445.D	12 Nov 2024 08:11		AR\AJ	Ok
55	P4602-28	CC0Q4	PS028446.D	12 Nov 2024 09:04		AR\AJ	Ok,M
56	I.BLK	I.BLK	PS028447.D	12 Nov 2024 09:28		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS111124

Review By	yogesh	Review On	11/12/2024 5:16:17 AM
Supervise By	Ankita	Supervise On	11/12/2024 9:00:51 AM
SubDirectory	PS111124	HP Acquire Method	HP Processing Method ps110924 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469
CCC Internal Standard/PEM	PP23462
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469

57	HSTDCCC750	HSTDCCC750	PS028448.D	12 Nov 2024 09:52		ARIAJ	Ok,M
----	------------	------------	------------	-------------------	--	-------	------

M : Manual Integration