

Surrogate Summary

SDG No.: **P4663**

Client: **Union Paving & Construction Company**

Analytical Method: **8151A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS028217.D	PIBLK-PS028217.D	2,4-DCAA	1	500	506	101		70 (39)	130 (175)
		2,4-DCAA	2	500	449	90		70 (39)	130 (175)
I.BLK-PS028224.D	PIBLK-PS028224.D	2,4-DCAA	1	500	486	97		70 (39)	130 (175)
		2,4-DCAA	2	500	559	112		70 (39)	130 (175)
P4663-02	TENNANT-PAD-2-3-STOCKPILE	2,4-DCAA	1	500	278	56	*	70 (10)	130 (141)
		2,4-DCAA	2	500	253	51	*	70 (10)	130 (141)
P4663-04	RES-BUILDING-PAD-NO-11	2,4-DCAA	1	500	269	54	*	70 (10)	130 (141)
		2,4-DCAA	2	500	247	49	*	70 (10)	130 (141)
P4663-06	RES-BUILDING-PAD-NO-3	2,4-DCAA	1	500	249	50	*	70 (10)	130 (141)
		2,4-DCAA	2	500	194	39	*	70 (10)	130 (141)
P4663-08	RES-BUILDING-PAD-NO-2	2,4-DCAA	1	500	359	72		70 (10)	130 (141)
		2,4-DCAA	2	500	330	66	*	70 (10)	130 (141)
PB164641BL	PB164641BL	2,4-DCAA	1	500	572	114		70 (10)	130 (141)
		2,4-DCAA	2	500	568	114		70 (10)	130 (141)
PB164641BS	PB164641BS	2,4-DCAA	1	500	573	115		70 (10)	130 (141)
		2,4-DCAA	2	500	572	114		70 (10)	130 (141)
I.BLK-PS028236.D	PIBLK-PS028236.D	2,4-DCAA	1	500	573	115		70 (39)	130 (175)
		2,4-DCAA	2	500	538	108		70 (39)	130 (175)
I.BLK-PS028355.D	PIBLK-PS028355.D	2,4-DCAA	1	500	501	100		70 (39)	130 (175)
		2,4-DCAA	2	500	503	101		70 (39)	130 (175)
I.BLK-PS028405.D	PIBLK-PS028405.D	2,4-DCAA	1	500	511	102		70 (39)	130 (175)
		2,4-DCAA	2	500	472	94		70 (39)	130 (175)
P4602-44MS	CC0R8MS	2,4-DCAA	1	500	375	75		70 (10)	130 (141)
		2,4-DCAA	2	500	253	51	*	70 (10)	130 (141)
P4602-45MSD	CC0R8MSD	2,4-DCAA	1	500	379	76		70 (10)	130 (141)
		2,4-DCAA	2	500	254	51	*	70 (10)	130 (141)
I.BLK-PS028417.D	PIBLK-PS028417.D	2,4-DCAA	1	500	513	103		70 (39)	130 (175)
		2,4-DCAA	2	500	468	94		70 (39)	130 (175)