

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4663

Client: Union Paving & Construction Company

Analytical Method: 8151A

DataFile : PS028408.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: P4602-44MS	CC0R8MS											
	DICAMBA	181.8	0	78.0	ug/Kg	43	*			70 (10)	130 (112)	
	DICHLORPROP	181.8	0	47.0	ug/Kg	26	*			70 (10)	130 (113)	
	2,4-D	181.8	0	90.4	ug/Kg	50	*			70 (10)	130 (144)	
	2,4,5-TP(Silvex)	181.8	0	29.7	ug/Kg	16	*			70 (10)	130 (114)	
	2,4,5-T	181.8	0	49.4	ug/Kg	27	*			70 (10)	130 (115)	
	2,4-DB	181.8	0	52.5	ug/Kg	29	*			70 (10)	130 (140)	
	Dinoseb	181.8	0	0	ug/Kg	0	*			70 (10)	130 (118)	