

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4663

Client: Union Paving & Construction Company

Analytical Method: 8151A

DataFile : PS028409.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	CC0R8MSD											
P4602-45MSD	DICAMBA	181.7	0	78.6	ug/Kg	43	*	0		70 (10)	130 (112)	30 (20)
	DICHLORPROP	181.7	0	47.4	ug/Kg	26	*	0		70 (10)	130 (113)	30 (20)
	2,4-D	181.7	0	91.1	ug/Kg	50	*	0		70 (10)	130 (144)	30 (20)
	2,4,5-TP(Silvex)	181.7	0	29.9	ug/Kg	16	*	0		70 (10)	130 (114)	30 (20)
	2,4,5-T	181.7	0	50.0	ug/Kg	28	*	4		70 (10)	130 (115)	30 (20)
	2,4-DB	181.7	0	52.5	ug/Kg	29	*	0		70 (10)	130 (140)	30 (20)
	Dinoseb	181.7	0	0	ug/Kg	0	*	0		70 (10)	130 (118)	30 (20)