

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4663

Client: Union Paving & Construction Company

Analytical Method: 8151A Datafile : PS028231.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB164641BS	DICAMBA	166.5	181	ug/Kg	109				70 (72)	130 (129)	
	DICHLORPROP	166.5	185	ug/Kg	111				70 (77)	130 (135)	
	2,4-D	166.5	186	ug/Kg	112				70 (65)	130 (144)	
	2,4,5-TP(Silvex)	166.5	194	ug/Kg	117				70 (74)	130 (146)	
	2,4,5-T	166.5	193	ug/Kg	116				70 (77)	130 (134)	
	2,4-DB	166.5	188	ug/Kg	113				70 (72)	130 (122)	
	Dinoseb	166.5	197	ug/Kg	118				70 (74)	130 (132)	