

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL01 Date Analyzed: 11/01/2024

Lab Sample No.: AR1660CCC500 Data File : PP068339.D Time Analyzed: 10:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.928	5.819	6.019	520.990	500.000	4.2
Aroclor-1016-2	5.950	5.842	6.042	531.040	500.000	6.2
Aroclor-1016-3	6.013	5.905	6.105	523.600	500.000	4.7
Aroclor-1016-4	6.112	6.003	6.203	521.790	500.000	4.4
Aroclor-1016-5	6.406	6.298	6.498	504.480	500.000	0.9
Aroclor-1260-1	7.530	7.422	7.622	484.880	500.000	-3.0
Aroclor-1260-2	7.782	7.675	7.875	480.060	500.000	-4.0
Aroclor-1260-3	8.143	8.037	8.237	490.870	500.000	-1.8
Aroclor-1260-4	8.382	8.275	8.475	488.400	500.000	-2.3
Aroclor-1260-5	8.720	8.612	8.812	490.380	500.000	-1.9
Decachlorobiphenyl	10.679	10.569	10.769	52.230	50.000	4.5
Tetrachloro-m-xylene	4.763	4.655	4.855	52.520	50.000	5.0

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL01 Date Analyzed: 11/01/2024

Lab Sample No.: AR1660CCC500 Data File : PP068339.D Time Analyzed: 10:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.161	5.061	5.261	464.570	500.000	-7.1
Aroclor-1016-2	5.181	5.081	5.281	469.790	500.000	-6.0
Aroclor-1016-3	5.362	5.262	5.462	473.350	500.000	-5.3
Aroclor-1016-4	5.401	5.301	5.501	479.830	500.000	-4.0
Aroclor-1016-5	5.620	5.520	5.720	480.720	500.000	-3.9
Aroclor-1260-1	6.662	6.564	6.764	461.980	500.000	-7.6
Aroclor-1260-2	6.849	6.750	6.950	462.820	500.000	-7.4
Aroclor-1260-3	7.007	6.908	7.108	464.260	500.000	-7.1
Aroclor-1260-4	7.480	7.383	7.583	468.850	500.000	-6.2
Aroclor-1260-5	7.719	7.622	7.822	465.090	500.000	-7.0
Decachlorobiphenyl	9.213	9.119	9.319	48.760	50.000	-2.5
Tetrachloro-m-xylene	4.054	3.952	4.152	49.060	50.000	-1.9