

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/01/2024

Lab Sample No.: AR1660CCC500 Data File : PP068354.D Time Analyzed: 18:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.925	5.819	6.019	514.430	500.000	2.9
Aroclor-1016-2	5.947	5.842	6.042	507.690	500.000	1.5
Aroclor-1016-3	6.011	5.905	6.105	499.770	500.000	0.0
Aroclor-1016-4	6.109	6.003	6.203	484.840	500.000	-3.0
Aroclor-1016-5	6.403	6.298	6.498	479.160	500.000	-4.2
Aroclor-1260-1	7.526	7.422	7.622	440.410	500.000	-11.9
Aroclor-1260-2	7.779	7.675	7.875	435.280	500.000	-12.9
Aroclor-1260-3	8.141	8.037	8.237	444.470	500.000	-11.1
Aroclor-1260-4	8.380	8.275	8.475	445.800	500.000	-10.8
Aroclor-1260-5	8.717	8.612	8.812	456.680	500.000	-8.7
Decachlorobiphenyl	10.676	10.569	10.769	48.090	50.000	-3.8
Tetrachloro-m-xylene	4.761	4.655	4.855	52.800	50.000	5.6

CALIBRATION VERIFICATION SUMMARY

Contract: UNIO02

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/01/2024

Lab Sample No.: AR1660CCC500 Data File : PP068354.D Time Analyzed: 18:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.158	5.061	5.261	476.910	500.000	-4.6
Aroclor-1016-2	5.178	5.081	5.281	473.500	500.000	-5.3
Aroclor-1016-3	5.359	5.262	5.462	494.850	500.000	-1.0
Aroclor-1016-4	5.398	5.301	5.501	492.210	500.000	-1.6
Aroclor-1016-5	5.616	5.520	5.720	484.560	500.000	-3.1
Aroclor-1260-1	6.659	6.564	6.764	446.980	500.000	-10.6
Aroclor-1260-2	6.846	6.750	6.950	447.550	500.000	-10.5
Aroclor-1260-3	7.003	6.908	7.108	442.460	500.000	-11.5
Aroclor-1260-4	7.478	7.383	7.583	434.030	500.000	-13.2
Aroclor-1260-5	7.716	7.622	7.822	429.410	500.000	-14.1
Decachlorobiphenyl	9.208	9.119	9.319	44.580	50.000	-10.8
Tetrachloro-m-xylene	4.051	3.952	4.152	49.330	50.000	-1.3