

Surrogate Summary

SDG No.: **P4663**

Client: **Union Paving & Construction Company**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP067586.D	PIBLK-PP067586.D	Tetrachloro-m-xylene	1	20	21.1	105		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.6	118		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	22.1	110		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	23.5	117		70 (60)	130 (140)
I.BLK-PP068343.D	PIBLK-PP068343.D	Tetrachloro-m-xylene	1	20	18.1	91		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.5	103		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	17.0	85		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	19.0	95		70 (60)	130 (140)
PB164590BL	PB164590BL	Tetrachloro-m-xylene	1	20	20.9	104		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	23.1	116		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.2	96		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.6	113		30 (32)	150 (175)
PB164590BS	PB164590BS	Tetrachloro-m-xylene	1	20	21.0	105		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	23.3	117		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.6	98		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.1	111		30 (32)	150 (175)
P4663-02	TENNANT-PAD-2-3-STOCKPILE	Tetrachloro-m-xylene	1	20	23.5	118		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	22.4	112		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.8	109		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.6	93		30 (32)	150 (175)
P4663-02MS	TENNANT-PAD-2-3-STOCKPILE	Tetrachloro-m-xylene	1	20	21.4	107		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.7	98		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.9	99		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	16.6	83		30 (32)	150 (175)
P4663-02MSD	TENNANT-PAD-2-3-STOCKPILE	Tetrachloro-m-xylene	1	20	21.6	108		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.2	96		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.2	101		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	16.5	82		30 (32)	150 (175)
P4663-04	RES-BUILDING-PAD-NO-11	Tetrachloro-m-xylene	1	20	22.8	114		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.8	99		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.3	106		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.0	90		30 (32)	150 (175)
P4663-06	RES-BUILDING-PAD-NO-3	Tetrachloro-m-xylene	1	20	22.5	113		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.7	99		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.2	106		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.1	90		30 (32)	150 (175)
P4663-08	RES-BUILDING-PAD-NO-2	Tetrachloro-m-xylene	1	20	18.5	92		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	16.9	85		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	17.2	86		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	15.5	78		30 (32)	150 (175)
I.BLK-PP068358.D	PIBLK-PP068358.D	Tetrachloro-m-xylene	1	20	19.1	95		70 (60)	130 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: P4663

Client: Union Paving & Construction Company

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP068358.D	PIBLK-PP068358.D	Decachlorobiphenyl	1	20	18.9	95		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	17.1	86		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	18.0	90		70 (60)	130 (140)