

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4663

Client: Union Paving & Construction Company

Analytical Method: 8081B

DataFile : PL092811.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Limits		
			Qual	RPD			Qual	Low	High	RPD		
Client Sample ID:	RES-BUILDING-PAD-NO-2MSD											
P4663-08MSD	alpha-BHC	18.23	0	15.5	ug/kg	85		0		30 (60)	150 (144)	30 (20)
	beta-BHC	18.23	0	17.8	ug/kg	98		1		30 (54)	150 (143)	30 (20)
	delta-BHC	18.23	0	4.60	ug/kg	25	*	0		30 (47)	150 (144)	30 (20)
	gamma-BHC (Lindane)	18.23	0	15.8	ug/kg	87		1		30 (61)	150 (140)	30 (20)
	Heptachlor	18.23	0	17.9	ug/kg	98		0		30 (63)	150 (135)	30 (20)
	Aldrin	18.23	0	17.0	ug/kg	93		1		30 (49)	150 (139)	30 (20)
	Heptachlor epoxide	18.23	0	17.4	ug/kg	95		2		30 (32)	150 (180)	30 (20)
	Endosulfan I	18.23	0	17.5	ug/kg	96		0		30 (56)	150 (142)	30 (20)
	Dieldrin	18.23	0	17.5	ug/kg	96		0		30 (47)	150 (161)	30 (20)
	4,4'-DDE	18.23	0	18.0	ug/kg	99		0		30 (55)	150 (136)	30 (20)
	Endrin	18.23	0	18.3	ug/kg	100		2		30 (57)	150 (139)	30 (20)
	Endosulfan II	18.23	0	17.3	ug/kg	95		3		30 (40)	150 (163)	30 (20)
	4,4'-DDD	18.23	0	17.4	ug/kg	95		5		30 (37)	150 (192)	30 (20)
	Endosulfan sulfate	18.23	0	13.7	ug/kg	75		3		30 (62)	150 (139)	30 (20)
	4,4'-DDT	18.23	0	16.8	ug/kg	92		0		30 (51)	150 (146)	30 (20)
	Methoxychlor	18.23	0	17.2	ug/kg	94		3		30 (54)	150 (136)	30 (20)
	Endrin ketone	18.23	0	15.9	ug/kg	87		0		30 (60)	150 (129)	30 (20)
	Endrin aldehyde	18.23	0	15.9	ug/kg	87		0		30 (59)	150 (132)	30 (20)
	alpha-Chlordane	18.23	0	17.7	ug/kg	97		1		30 (30)	150 (192)	30 (20)
	gamma-Chlordane	18.23	0	17.9	ug/kg	98		0		30 (44)	150 (175)	30 (20)