

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4663

Client: Union Paving & Construction Company

Analytical Method: **8081B**

Datafile : PL092802.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB164592BS	alpha-BHC	16.65	17.7	ug/kg	106				40 (84)	140 (123)	
	beta-BHC	16.65	18.0	ug/kg	108				40 (82)	140 (123)	
	delta-BHC	16.65	15.6	ug/kg	94				40 (83)	140 (126)	
	gamma-BHC (Lindane)	16.65	17.6	ug/kg	106				40 (83)	140 (125)	
	Heptachlor	16.65	18.7	ug/kg	112				40 (83)	140 (122)	
	Aldrin	16.65	17.8	ug/kg	107				40 (82)	140 (124)	
	Heptachlor epoxide	16.65	18.4	ug/kg	111				40 (83)	140 (120)	
	Endosulfan I	16.65	18.7	ug/kg	112				40 (81)	140 (124)	
	Dieldrin	16.65	18.9	ug/kg	114				40 (85)	140 (121)	
	4,4'-DDE	16.65	18.8	ug/kg	113				40 (81)	140 (123)	
	Endrin	16.65	19.8	ug/kg	119				40 (76)	140 (130)	
	Endosulfan II	16.65	19.2	ug/kg	115				40 (80)	140 (125)	
	4,4'-DDD	16.65	19.0	ug/kg	114				40 (80)	140 (131)	
	Endosulfan sulfate	16.65	18.6	ug/kg	112				40 (81)	140 (122)	
	4,4'-DDT	16.65	19.3	ug/kg	116				40 (70)	140 (129)	
	Methoxychlor	16.65	18.7	ug/kg	112				40 (60)	140 (119)	
	Endrin ketone	16.65	18.7	ug/kg	112				40 (77)	140 (132)	
	Endrin aldehyde	16.65	18.1	ug/kg	109				40 (79)	140 (124)	
	alpha-Chlordane	16.65	18.7	ug/kg	112				40 (84)	140 (120)	
	gamma-Chlordane	16.65	18.9	ug/kg	114				40 (83)	140 (122)	