

rteres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG110124
Data File : FG014632.D
Signal(s) : FID1A.ch
Acq On : 01 Nov 2024 21:22
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/04/2024
Supervised By :Ankita Jodhani 11/04/2024

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG102124.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.064	2.030	2.128	PV	554880	6143515	82.39%	4.929%
2	4.581	4.509	4.645	BB	708109	7259204	97.35%	5.824%
3	6.753	6.708	6.808	BV	729483	7351063	98.59%	5.898%
4	8.584	8.536	8.628	BB	690175	7358523	98.69%	5.904%
5	10.196	10.152	10.253	BB	652567	7355476	98.65%	5.901%
6	11.642	11.526	11.696	BV	642043	7438886	99.76%	5.968%
7	12.955	12.833	13.018	BB	623210	7456496	100.00%	5.982%
8	14.156	14.118	14.228	VB	572085	7161442	96.04%	5.745%
9	15.056	15.013	15.084	BV	464701	6322305	84.79%	5.072%
10	15.261	15.189	15.310	VB	569605	7071329	94.83%	5.673%
11	16.282	16.188	16.329	BV	533262	6942593	93.11%	5.570%
12	17.233	17.145	17.299	BV	531855	6951064	93.22%	5.577%
13	18.122	18.037	18.174	VV	525719	7329928	98.30%	5.881%
14	18.956	18.885	18.997	PV	522567	7395987	99.19%	5.934%
15	19.741	19.700	19.802	BV	470161	6979145	93.60%	5.599%
16	20.479	20.418	20.517	VV	418056	6423052	86.14%	5.153%
17	21.231	21.174	21.300	PV	319720	5974132	80.12%	4.793%
18	22.176	22.095	22.285	VV	229242	5732102	76.87%	4.599%
Sum of corrected areas:						124646240		

FG102124.M Mon Nov 04 02:32:56 2024