

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: UNIO02
 ProjectID: Rt. 10 Whippany
 Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG No.: P4663

Calibration Sequence : FG102124		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	202059326	118858	FG014453.D
850	102753431	120886	FG014454.D
340	42726636	125667	FG014455.D
170	22264935	130970	FG014456.D
85	11318543	133159	FG014457.D
AVG RF : 125908		% RSD : 4.915	AVG RT : 15.0822