

SOIL TPH GC SURROGATE RECOVERY

Lab Name: Chemtech Client: Union Paving & Construction Company
Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG No.: P4663

EPA SAMPLE NO.	S1 TETRACOSANE-d50	S2	S3	S4	TOT OUT
PIBLK-FG014617.D	89				0
PIBLK-FG014625.D	88				0
PIBLK-FG014631.D	97				0
TENNANT-PAD-2-3-STOCKPILE	54				0
RES-BUILDING-PAD-NO-11	67				0
RES-BUILDING-PAD-NO-3	67				0
RES-BUILDING-PAD-NO-2	75				0
RES-BUILDING-PAD-NO-2MS	62				0
RES-BUILDING-PAD-NO-2MSD	59				0
PB164602BL	80				0
PB164602BS	82				0

QC LIMITS

TETRACOSANE-d50

For Water : 29-130

For Soil : 37-130

Column to be used to flag recovery values
* Values outside of contract required QC limits
D Surrogate Diluted Out