

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: UNIO02
Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG No.: P4663
Instrument ID: BNA_F Calibration Date(s): 10/18/2024 10/18/2024
Calibration Time(s): 10:27 13:46

LAB FILE ID:		RRF2.5 = BF139844.D			RRF005 = BF139845.D		RRF010 = BF139846.D	
		RRF020 = BF139847.D			RRF040 = BF139848.D		RRF050 = BF139849.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.465	1.398	1.331	1.278	1.179	1.278	10.1
Benzaldehyde			1.137	1.042	0.940	0.873	0.931	15.2
Phenol-d6		1.900	1.818	1.709	1.647	1.538	1.655	10.1
Phenol		1.952	1.832	1.760	1.712	1.583	1.706	9.2
bis(2-Chloroethyl) ether		1.470	1.423	1.359	1.294	1.251	1.319	7.8
2-Chlorophenol		1.503	1.422	1.358	1.310	1.212	1.306	10.2
2-Methylphenol		1.264	1.164	1.134	1.114	1.053	1.114	7.7
2,2-oxybis(1-Chloropropane)		2.524	2.409	2.353	2.255	2.117	2.248	8.7
Acetophenone		0.578	0.533	0.516	0.481	0.452	0.490	11.4
3+4-Methylphenols		1.678	1.573	1.520	1.418	1.309	1.424	12.4
n-Nitroso-di-n-propylamine	1.105	1.141	1.066	1.025	0.974	0.921	1.004	9.6
Nitrobenzene-d5		0.365	0.365	0.372	0.371	0.354	0.361	2.9
Hexachloroethane		0.583	0.571	0.549	0.541	0.507	0.534	7.1
Nitrobenzene		0.417	0.402	0.408	0.403	0.383	0.396	4.1
Isophorone		0.755	0.709	0.702	0.679	0.652	0.685	5.9
2-Nitrophenol		0.124	0.137	0.150	0.157	0.158	0.149	9.2
2,4-Dimethylphenol		0.285	0.259	0.256	0.248	0.237	0.249	8.1
bis(2-Chloroethoxy) methane		0.468	0.440	0.432	0.412	0.394	0.415	8.2
2,4-Dichlorophenol		0.308	0.297	0.293	0.283	0.272	0.283	6.2
Naphthalene		1.209	1.121	1.091	1.023	0.958	1.031	11.2
4-Chloroaniline		0.397	0.382	0.368	0.351	0.332	0.352	9.3
Hexachlorobutadiene		0.224	0.206	0.203	0.197	0.185	0.197	8.0
Caprolactam		0.092	0.092	0.092	0.092	0.088	0.090	2.9
4-Chloro-3-methylphenol		0.341	0.328	0.324	0.315	0.299	0.314	6.0
2-Methylnaphthalene		0.740	0.689	0.666	0.621	0.587	0.632	11.1
Hexachlorocyclopentadiene		0.185	0.193	0.198	0.198	0.186	0.188	5.3
2,4,6-Trichlorophenol		0.405	0.382	0.400	0.387	0.363	0.382	4.8
2-Fluorobiphenyl		1.512	1.396	1.280	1.162	1.088	1.210	16.0
2,4,5-Trichlorophenol		0.426	0.425	0.398	0.401	0.381	0.394	6.6
1,1-Biphenyl		1.640	1.536	1.483	1.379	1.285	1.392	12.2
2-Chloronaphthalene		1.288	1.225	1.164	1.111	1.048	1.121	10.0
2-Nitroaniline		0.299	0.318	0.353	0.365	0.354	0.343	7.2
Dimethylphthalate		1.410	1.344	1.283	1.237	1.172	1.246	8.6
Acenaphthylene		1.854	1.756	1.717	1.615	1.507	1.621	10.1

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: UNIO02
Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG No.: P4663
Instrument ID: BNA_F Calibration Date(s): 10/18/2024 10/18/2024
Calibration Time(s): 10:27 13:46

LAB FILE ID:		RRF2.5 = BF139844.D			RRF005 = BF139845.D		RRF010 = BF139846.D	
		RRF020 = BF139847.D			RRF040 = BF139848.D		RRF050 = BF139849.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.256	0.266	0.278	0.280	0.273	0.270	3.4
3-Nitroaniline		0.270	0.275	0.296	0.292	0.276	0.278	4.8
Acenaphthene		1.200	1.136	1.089	1.044	0.977	1.049	9.5
2,4-Dinitrophenol			0.056	0.077	0.101	0.101	0.093	23.9
4-Nitrophenol		0.187	0.207	0.225	0.232	0.219	0.214	6.8
Dibenzofuran		1.767	1.659	1.579	1.486	1.389	1.505	11.5
2,4-Dinitrotoluene		0.280	0.314	0.337	0.356	0.344	0.332	7.9
Diethylphthalate		1.366	1.308	1.267	1.214	1.165	1.223	8.0
4-Chlorophenyl-phenylether		0.698	0.638	0.617	0.569	0.526	0.579	13.1
Fluorene		1.409	1.309	1.201	1.110	1.029	1.146	14.7
4-Nitroaniline		0.249	0.259	0.270	0.275	0.263	0.263	3.6
4,6-Dinitro-2-methylphenol			0.055	0.072	0.087	0.090	0.082	18.6
n-Nitrosodiphenylamine		0.672	0.641	0.621	0.595	0.568	0.599	8.1
2,4,6-Tribromophenol		0.201	0.196	0.193	0.188	0.179	0.187	5.5
4-Bromophenyl-phenylether		0.230	0.217	0.211	0.202	0.197	0.206	7.0
Hexachlorobenzene		0.258	0.245	0.237	0.231	0.217	0.232	7.2
Atrazine		0.189	0.175	0.156	0.175	0.135	0.162	11.4
Pentachlorophenol		0.118	0.137	0.148	0.152	0.145	0.141	8.0
Phenanthrene		1.121	1.035	0.994	0.933	0.863	0.945	11.8
Anthracene		1.082	1.014	0.972	0.913	0.851	0.922	11.5
Carbazole		1.002	0.964	0.923	0.846	0.776	0.857	12.6
Di-n-butylphthalate		1.104	1.071	1.062	1.014	0.923	0.990	9.8
Fluoranthene		1.149	1.108	1.036	0.943	0.842	0.955	15.3
Pyrene		1.892	1.828	1.900	1.805	1.685	1.751	8.4
Terphenyl-d14		1.381	1.335	1.340	1.244	1.154	1.227	11.0
Butylbenzylphthalate		0.490	0.513	0.536	0.555	0.535	0.525	4.0
3,3-Dichlorobenzidine		0.368	0.372	0.390	0.387	0.374	0.380	2.2
Benzo (a) anthracene		1.425	1.355	1.329	1.331	1.267	1.302	6.5
Chrysene		1.325	1.244	1.234	1.167	1.134	1.194	6.5
Bis (2-ethylhexyl) phthalate		0.520	0.539	0.577	0.626	0.620	0.589	7.5
Di-n-octyl phthalate			0.786	0.930	1.135	1.182	1.071	16.1
Benzo (b) fluoranthene		1.317	1.240	1.319	1.196	1.105	1.218	7.1
Benzo (k) fluoranthene		1.213	1.177	0.992	1.066	1.030	1.051	10.4
Benzo (a) pyrene		1.030	1.024	1.018	1.025	0.974	1.002	3.1
Indeno (1,2,3-cd) pyrene		1.209	1.261	1.307	1.352	1.299	1.287	3.8

All other compounds must meet a minimum RRF of 0.010.

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: UNIO02
Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG No.: P4663
Instrument ID: BNA_F Calibration Date(s): 10/18/2024 10/18/2024
Calibration Time(s): 10:27 13:46

LAB FILE ID:		RRF2.5 = BF139844.D		RRF005 = BF139845.D		RRF010 = BF139846.D		
		RRF020 = BF139847.D		RRF040 = BF139848.D		RRF050 = BF139849.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		1.021	1.064	1.103	1.120	1.083	1.073	3.3
Benzo (g,h,i) perylene		1.030	1.046	1.090	1.128	1.081	1.072	3.4
1,2,4,5-Tetrachlorobenzene		0.609	0.579	0.562	0.537	0.512	0.540	8.7
1,4-Dioxane		0.651	0.625	0.603	0.591	0.571	0.596	5.7
2,3,4,6-Tetrachlorophenol		0.334	0.322	0.326	0.310	0.296	0.309	6.3