

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4663 SAS No.: P4663 SDG NO.: P4663

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/01/2024

Lab File ID: BE101444.D Time Analyzed: 09:41

Instrument ID: BNA_E GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	92171	7.576	414792	10.34	283187	14.19
UPPER LIMIT	184342	8.076	829584	10.843	566374	14.686
LOWER LIMIT	46085.5	7.076	207396	9.843	141594	13.686
EPA SAMPLE NO.						
01 TENNANT-PAD-2-3-STOCKPILE	90671	7.57	396894	10.34	265226	14.18
02 RES-BUILDING-PAD-NO-3	69420	7.58	296978	10.34	197789	14.18

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.